



Development and evaluation of a suite of isotope reference gases for methane in air

5 P. Sperlich^{1, 2}, N. A. M. Uitslag^{1,3}, J. M. Richter¹, M. Rothe¹, H. Geilmann¹, C. van der Veen⁴, T. Röckmann⁴, T. Blunier⁵ and W. A. Brand¹

¹Max-Planck-Institute for Biogeochemistry (MPI-BGC), Jena, Germany

²National Institute of Water and Atmospheric Research (NIWA), Wellington, New Zealand

10 ³Centre for Isotope Research (CIO), Energy and Sustainability Research Institute Groningen (ESRIG), University of Groningen, Groningen, The Netherlands

⁴Institute of Marine and Atmospheric Science in Utrecht (IMAU), Utrecht, The Netherlands

⁵Centre for Ice and Climate (CIC), Niels Bohr Institute, University of Copenhagen, Copenhagen, Denmark

Correspondence to: P. Sperlich (peter.sperlich@niwa.co.nz)

15 **Abstract.** Measurements made by multiple analytical facilities can only be comparable if they are related to a unifying and traceable reference. However, reference materials that fulfil these fundamental requirements are unavailable for the analysis of isotope ratios in atmospheric methane, which led to misinterpretations of combined data sets in the past. We developed a method to produce a suite of standard gases that can be used to unify methane isotope ratio measurements of laboratories in the atmospheric monitoring community. We calibrated a suite of pure methane gases of different methanogenic origin against
20 international referencing materials that define the VSMOW and VPDB isotope scales. The isotope ratios of our pure methane gases range between -320 and $+40$ ‰ for $\delta^2\text{H-CH}_4$ and between -70 and -40 ‰ for $\delta^{13}\text{C-CH}_4$, enveloping the isotope ratios of tropospheric methane (about -90 ‰ and -47 ‰ for $\delta^2\text{H-CH}_4$ and $\delta^{13}\text{C-CH}_4$, respectively). We estimate combined uncertainties for our $\delta^2\text{H}$ and $\delta^{13}\text{C}$ calibrations of <1.5 ‰ and <0.2 ‰, respectively.

25 Aliquots of the calibrated pure methane gases have been diluted with methane-free air to atmospheric methane levels and filled into 5-L glass flasks. These synthetic gas mixtures comprise atmospheric oxygen/nitrogen ratios as well as appropriate argon, krypton and nitrous oxide mole fractions to prevent gas-specific measurement artefacts. The resulting synthetic atmospheric reference gases will be available to the atmospheric monitoring community. This will provide unifying isotope scale anchors for isotope ratio measurements of atmospheric methane so that data sets can be merged into a consistent global data frame.



1 Introduction

Isotope ratios of trace gases in the present and the past atmosphere (e.g. from ice cores) are an emerging tool to study the biogeochemical processes that cause the variation of CH₄ in the atmosphere (Stevens and Rust, (1982), Lowe et al., (1994), Sapart et al., (2012), Sperlich et al., (2015)). Recently, two conflicting publications highlighted *i*) the interpretative power when data sets from multiple laboratories are combined for spatio-temporal analysis of CH₄ isotope ratios (Kai et al., 2011) and *ii*) the pitfalls when differences due to laboratory offsets are mis-interpreted as spatial variability of CH₄ sources (Levin et al., 2012). Levin et al., (2012) identified calibration offsets between three laboratories by comparing their long term observations in Antarctica where the carbon isotopic composition of CH₄ is assumed to be free of spatial gradients. However, this technique is a temporary work-around that excludes the use of data sets from laboratories without a history of observations in Antarctica or a traceable link to Antarctic observations. If a unique set of reference material were available to all laboratories that measure isotope ratios of atmospheric CH₄, the science community would be able to combine the available measured data sets to investigate spatio-temporal variations. However, such a sustainable solution that unifies the measurements of all laboratories has not been established.

The lack of unique reference gases has long been recognised in the literature on CH₄ isotope ratios, ranging from pioneering papers (e.g. Craig (1953), Schiegl and Vogel, (1970)) to recent technical publications on isotope ratios in atmospheric CH₄ (e.g. Sperlich et al., (2013), Tokida et al., (2014), Eyer et al., (2015)) as well as papers that present and interpret such data (Levin et al., 2012). So far, laboratories have either developed methods to calibrate purified CH₄ against international reference material of different physico-chemical properties (e.g. Schiegl and Vogel, (1970), Stevens and Rust, (1982), Dumke et al., (1989), Levin et al., (1993), Lowe et al., (1994), Quay et al., (1999), Sperlich et al., (2012)) or they reference their CH₄ samples to standards that were propagated from calibrations elsewhere (e.g. Bergamaschi et al., (1994), Brass and Röckmann, (2010), Bock et al., (2014), Schmitt et al., (2014), Rella et al., (2015), W.A. Brand 2015 pers. comm.). The diversity of referencing trajectories can lead to significant referencing offsets between the laboratories as shown by Levin et al., (2012).

Ghosh et al., (2005) established a method to harmonise isotope ratio measurements of atmospheric CO₂. Based on this method, the ISOLAB of the Max-Planck-Institute for Biogeochemistry (MPI-BGC) in Jena, Germany, distributes a suite of reference gases, known as the “JRAS air set”, which is accepted as an isotope scale anchor by the community (WMO, 2012). Calibrating against the “JRAS air set” reduces laboratory offsets and helps reaching and maintaining the compatibility goal (Wendeberg et al., 2013).

This paper describes a method to produce isotope reference gases for atmospheric CH₄. The reference gases we produced bracket the isotopic ratios of tropospheric CH₄ and span over a large isotopic abundance range, which enables a two-point calibration to account for scale compression effects, thereby reducing the referencing uncertainty (Coplen et al., 2006). These isotope reference gases are produced from a well calibrated suite of CH₄ gases and will be made available from MPI-BGC. Our reference gases for CH₄ isotope ratios may help the community to reach the compatibility goals of 1 ‰ and 0.02 ‰ for δ²H-CH₄ and δ¹³C-CH₄, respectively (WMO, 2014).



2 Materials and Methods

The aim of this paper is to present the method that we use to calibrate and prepare isotope reference gases for CH₄ in air samples, as outlined in the flow diagram of Fig. 1. Our method is based on the initial calibration of two pure Master-CH₄ gases for their $\delta^2\text{H-CH}_4$ and $\delta^{13}\text{C-CH}_4$ isotope ratios against international isotope reference materials. The two Master-CH₄ gases are then used to calibrate a suite of further CH₄ gases. Once calibrated, aliquots of the pure CH₄ gases are diluted with CH₄-free air to atmospheric CH₄ mixing ratios. The resulting synthetic air standard can be distributed and analysed in a similar fashion as atmospheric samples by collaborating laboratories, following the Principle of Identical Treatment, PIT, (Werner and Brand, 2001).




Figure 1: Flow diagram of methodological steps to produce synthetic isotope reference gases for CH₄ in air. RM is the abbreviation for established international Reference Materials.

2.1 Gases and reference materials used in this study

Our study is based on a suite of CH₄ gases that differ in their commercial provider, methanogenic origin and isotopic composition. We identify our CH₄ gases by names as shown in Table 1. “Biogenic” and “Fossil” have been calibrated in a previous CH₄ referencing study (Sperlich et al., 2012) and are therefore of known isotopic composition. Treating these gases as unknowns in this study allows testing and evaluating the performance of the presented methods. Six other CH₄ gases were purchased from suppliers of commercial gases or laboratory equipment (Air-Liquide, Westfalen AG, Linde, Messer, Campro Scientific) and were used as purchased or as mixtures thereof. The purity level of all our CH₄ gases is 99.995% or higher.



Table 1: Gases used for this study. Note that Mike-1 and Martha-1 were created as gas mixtures but were topped up with further gases to create Mike-2 and Martha-2, respectively. Thus, Mike-1 and Martha-1 do not exist anymore.

Gas name	Cylinder volume [L]	pressure [bar]	function in study	CH ₄ source	gas supplier
Megan	10	lost	1 st Master CH ₄	fossil CH ₄	Air Liquide
Merlin	10	190	2 nd Master CH ₄	fossil CH ₄	Air Liquide
Mike-1	-	-	calibration CH ₄	MPI mixture	
Mike-2	5	45	calibration CH ₄	MPI mixture	
Merkur	2	100	calibration CH ₄	fossil CH ₄	Messer Griesheim
Merida	10	175	calibration CH ₄	Unknown	Westfalen AG
Martha-1	-	-	calibration CH ₄	MPI mixture	
Martha-2	10	165	calibration CH ₄	MPI mixture	
Minion	3	150	calibration CH ₄	Unknown	Messer Griesheim
Melly	50	193	calibration CH ₄	Unknown	Westfalen AG
δ ² H-Spike gas	0.4	2.5	spiking gas CH ₃ D		Campro Scientific
Fossil	30	2	calibration control	fossil CH ₄	Air Liquide Denmark
Biogenic	30	2	calibration control	biogas plant	Biogas Plant
synthetic air	50	200	synthetic air matrix		Linde
Krypton	2	200	synthetic air matrix		Westfalen AG
Carina-1	50	200	isotope scale	Jena air	MPI-BGC
Carina-2	50	200	isotope scale	Jena air	MPI-BGC

5

“Megan” and “Merlin” were used as Master-CH₄ gases that were calibrated against international reference materials for their hydrogen and carbon isotopic composition. Table 2 summarises information on the applied reference materials.

10

Table 2: Reference materials used in this study. Column 1 and 2 show name and chemical formula, respectively. “RM” in column 3 indicates an international reference material while “ws” identifies local working standards. The δ²H and δ¹³C isotope ratios are shown in column 4 and 5, respectively, while column 6 and 7 list the sources and related publications of reference materials we used. We show the most recent values of reference materials as summarized by Brand et al., (2014). For MPI-BGC data the uncertainties correspond to the 95 % confidence limit of the error of the mean, i.e. the standard error of the mean, multiplied by Student’s t-factor. The * indicates an uncertainty estimate that is currently under investigation as explained in the main text.



Name	Material	int/ws	$\delta^2\text{H}$ [‰]	$\delta^{13}\text{C}$ [‰]	Source	Reference
VSMOW-2	H ₂ O	RM	0 ± 0.3	--	IAEA	Gröning et al., (2007)
SLAP-2	H ₂ O	RM	-427.5 ± 0.3	--	IAEA	Gröning et al., (2007)
GISP	H ₂ O	RM	-189.7 ± 0.9	--	IAEA	Brand et al., (2014)
LSVEC	Li ₂ CO ₃	RM	--	-46.6 ± 0.03*	IAEA	Coplen et al., (2006),
CO-9	BaCO ₃	RM	--	-47.32 ± 0.06	IAEA	Coplen et al., (2006)
www-j1	H ₂ O	ws	-67.0 ± 0.4	--	MPI-BGC	--
BGP-j1	H ₂ O	ws	-187.1 ± 0.6	--	MPI-BGC	--
Mar-j1	CaCO ₃	ws	--	1.96 ± 0.01	MPI-BGC	Brand et al., (2009b)
ali-j3	Acetanilide	ws	--	-30.06 ± 0.05	MPI-BGC	--

2.2 Referencing CH₄ for $\delta^2\text{H}$ against VSMOW/SLAP and against other CH₄ gases

We use a high temperature conversion elemental analyser (TC/EA) coupled to an IRMS (Delta Plus XL, Thermo Finnigan, Bremen, Germany) via an open split (ConFlo III, Thermo Finnigan, Bremen, Germany) to convert CH₄ to H₂ (+ carbon) and measure $\delta^2\text{H}$ -CH₄ in pure CH₄ gases after the conversion to H₂ (Werner et al., 1999). This system is routinely used for the analysis of hydrogen and oxygen isotope ratios in water samples that are injected through a heated septum at (130°C) into a glassy carbon reactor where the H₂O is converted to CO and H₂ at temperatures above 1400°C (Gehre et al., (2004), Hilker et al., (1999)). In order to reference the hydrogen isotopic composition of pure CH₄ against international reference waters (VSMOW-SLAP scale), we inject pure CH₄ into the TC/EA using a two position 10-port valve (VICI, USA) that is configured as shown in Fig. 2. The outlet flow of this valve is routed through the septum also used for water injection so that the helium gas stream of 15 mL/min carries the CH₄ from the 1 mL sample loops into the TC/EA reactor. Typical CH₄ feed flow rates range between 2-3 mL/minute. The upstream plumbing can be configured so that every valve switch injects an aliquot of the same or of two alternating CH₄ gases into the reactor. Measurement sequences can now be created by injecting CH₄ gases manually and reference waters automatically (via injections through the septum from an autosampler) into the reactor, where both CH₄ and H₂O are converted to molecular hydrogen. Apart from the injection procedure itself, the generated H₂ of both CH₄ and international reference waters conforms with the Principle of Identical analytical Treatment (PIT) (Werner and Brand, 2001). The amounts of injected CH₄ and H₂O are adjusted in our experiments to achieve matching peak shapes and amplitudes during IRMS analysis in order to minimize H₃-factor variations between CH₄ and H₂O samples. The resulting peaks for CH₄ samples and reference waters are shown in Fig. 4. Following this method, we calibrated “Megan” and “Merlin” (Table 1) in three independent repetitions during three days against the in house standards “www-j1” and “BGP-j1”, which were independently calibrated against international reference waters VSMOW2 and SLAP2 (Table 2).




Figure 2: Configuration of two-position 10-port valve with plumbing for helium carrier gas, two ports to introduce the CH₄ sample and two 1-mL sample loops as indicated by thick blue lines. Note that vent 1 can be connected to CH₄ port 2 in order to introduce the same CH₄ gas with every injection. Sample loop 1 is vented through vent 1 for the injection of two alternating
5 CH₄ gases.

The two sample loops attached to the 10-port valve (Fig. 2) can be fed by two different CH₄ gases. This configuration enables alternating measurements of unknown CH₄ against known CH₄ in an identical fashion. When used in this configuration, the
10 previously calibrated “Megan” and “Merlin” served as known reference gases to calibrate the other CH₄ gases (Table 1) to the respective isotope scales. Additional injections of H₂O standards were used to control isotope scale contraction during $\delta^2\text{H}$ -CH₄ analysis.

Because the $\delta^2\text{H}$ -CH₄ range in our samples is very large (~350 ‰, Table 3), sufficient control of potential scale compression effects is imperative (Renpenning et al., 2015). We achieve this by applying reference waters that span the full VSMOW2 to
15 SLAP2 isotopic range from 0 to -427.5 ‰. The daily working reference waters (www-J1 and BGP-J1) cover a $\delta^2\text{H}$ range of -67.0 to -187.1 ‰ (Table 2). Our suite of reference waters enables a multi-point calibration that largely envelopes the $\delta^2\text{H}$ range of the CH₄ samples (Table 2 and 3).

In most routine isotope ratio measurements, samples are referenced against a standard of identical composition, following the principle of identical treatment (Werner and Brand, 2001). For example, measurement artefacts cancel when water samples
20 are calibrated against reference waters. However, great care has to be taken when comparing two different materials, as is the case when calibrating CH₄ against reference waters. This is because standard and sample require material-specific attention to prevent isotope fractionation during analysis. Calibration errors may occur when *i*) only one of the two materials or *ii*) both



materials are incompletely transported or converted, or when memory effects inside the sample conversion system are different for the respective materials. In order to rule out calibration artefacts due to the analysis of two chemically very different materials and in order to ensure that our method is suitable for quantitative analysis of both CH₄ and H₂O without isotope fractionation, we performed several series of experiments to optimise the reactor and septum temperatures, the compound-specific memory effects and the applied sample size as described in detail in Appendix 1.

The introduction of H₂ samples into the ion source of an IRMS leads to the formation of H₃⁺ ions that are registered on the HD⁺ detector, which is accounted for by the so called “H₃-factor correction” (Friedman, (1953), Sessions et al., (2001)). While the H₃-factor correction is experimentally determined and kept constant throughout a series of measurements, the H₃⁺ formation may be dynamic over time. In order to minimise the impact of inaccurate H₃-correction between CH₄ and H₂O analysis, we matched the H₂ peak heights resulting from both CH₄ and H₂O injections around 5.5 ± 0.5 V. Typical peak widths were around 45 s and 60 s for H₂O- and CH₄-derived H₂ peaks, respectively. A typical chromatogram of the δ²H-CH₄ calibration including details on peak shape and background is shown in Fig. 3. This figure shows that both the CH₄-derived and the H₂O-derived H₂ peaks are of comparable height and width, which justifies the analysis of sample and reference material using the standard integration software (ISODAT, Thermo). We performed a range of experiments to test for systematic analytical effects that could impact on the isotope analysis of CH₄ and/or H₂O and used the results of these experiments to optimize the analysis. These experiments are described in detail in Appendix 1. The good control of peak shapes and the elimination of analytical artefacts allow for the conclusion that our method is suitable to calibrate CH₄ against reference H₂O with high precision and accuracy.




Figure 3: Chromatograms of $\delta^2\text{H-CH}_4$ calibration sequences using TC/EA-IRMS with traces of m/z 2 and m/z 3 shown in black and blue, respectively. The bottom panel shows an example of an entire calibration sequence which begins with 3 square-shaped peaks of pure H₂, followed by alternations of 3 to 4 H₂O- and 3 to 4 CH₄-derived H₂ peaks before the sequence ends with another 3 square shaped peaks of pure H₂. The top left panel enlarges 4 peaks that are derived from H₂O (peak # 6-7) and CH₄ (peak # 8-9), respectively. A zoom into baseline details of H₂O-derived peak # 7 and CH₄-derived peak # 8 is shown in the top right panel. Red lines indicate the sections used for peak integration (peak widths are 43 s and 59 s for H₂O- and CH₄-derived H₂ peaks) by the IRMS software.


10

15



2.3 Referencing CH₄ against LSVEC / NBS 19

We calibrated $\delta^{13}\text{C}\text{-CH}_4$ in pure CH₄ gases after conversion to CO₂ using an elemental analyser (EA 1100, CE, Rodano, Italy) coupled to an IRMS (Delta Plus, Thermo Finnigan, Bremen, Germany) through an open split (ConFlo III, Thermo Finnigan, Bremen, Germany). This system is routinely used for the analysis of ¹³C and ¹⁵N in samples with solid or liquid matrices (Werner et al., (1999), Brooks et al., (2003)). For the CH₄ calibrations, we used the same 10-port valve assembly as described above (Fig. 2). We modified the EA autosampler to inject the CH₄ gas with a helium carrier gas stream of 10 mL/min through an extra 1/16" tube of 70/30 % Cu/Ni alloy into the oxygen plume region inside the combustion chamber of the EA (Fig. 4).




10 Figure 4: EA-IRMS system with additional inlet for gas injection into the combustion unit.

The plumbing of the system is designed so that gaseous and solid samples can be supplied to the same location inside the combustion reactor of the EA. All samples are oxidised at 1020°C (Werner et al., 1999) and experience identical analytical treatment after the combustion. After oxidation, the sample passes through a reduction reactor filled with elemental copper at 650°C to remove excess O₂ and to reduce NO_x if present. The sample is dried using a combination of a Nafion™ dryer (Perma Pure LLC, Toms River, NJ, USA, not shown in Fig 4) and Mg(ClO₄)₂ before it is routed through a GC column that is held at 80°C (3m, 1/4", Porapak PQS, CE instruments). The sample is then transferred into the IRMS through an open split.



Alternating CH₄ injections and applications of reference materials such as LSVEC, Mar-j1 and ali-j3 (Table 2) via an autosampler enable direct referencing of CH₄ gas to the VPDB isotope scale. We used this method to calibrate “Megan” and “Merlin” as Master-CH₄ gases (Table 1) over three independent measurement periods each. Alternating injections of a Master-CH₄ and unknown CH₄ gases were then used to calibrate all other CH₄ gases (Table 1) for $\delta^{13}\text{C}$. Two chromatograms resulting from both CH₄ and carbonate analysis using EA-IRMS are displayed in Fig. 5 and show very similar peak shapes for the analysis of both materials. Typical m/z 44 amplitudes and peak widths were $\sim 7.4 \pm 0.2$ V and 101 ± 1 s for both materials, respectively.



10 Figure 5: Chromatograms of $\delta^{13}\text{C}$ -CH₄ calibrations using EA-IRMS with traces for m/z 44, 45 and 46 represented by green, brown and black lines, respectively. Analysis of CH₄ (Merlin) and Li₂CO₃ (LSVEC) are shown in the left and right-hand panels, respectively. Bottom panels show the entire chromatogram while the two top panels zoom into the baseline of the traces. The first three square-shaped CO₂ peaks stem from injections of pure CO₂ working gas while the tailing peaks result



from CH₄- and Li₂CO₃-derived CO₂ analysis. The two red lines in the top panel indicate the sections that the IRMS software uses for peak integration (101 s and 100 s for CH₄- and Li₂CO₃-derived CO₂ peaks, respectively).

5 2.4 Measurement uncertainty and error propagation

With the two-step calibration strategy approach the error treatment also has two major components. For the $\delta^{13}\text{C}_{\text{VPDB}}$ calibration of the two Master-CH₄ gases Megan and Merlin we used the international reference materials NBS 19 and LSVEC with assigned $\delta^{13}\text{C}_{\text{VPDB}}$ values +1.95 and -46.6 ‰, respectively (Coplen et al., (2006), Brand et al., (2014)). These were analysed within the same analytical sequences as the Master-CH₄ gases and used to position the $\delta^{13}\text{C}$ -CH₄ values onto the international VPDB scale within every sequence by relating the $\delta^{13}\text{C}$ -CH₄ value of each CH₄ peak to the known $\delta^{13}\text{C}$ values of the reference materials. The final uncertainty for the Master-CH₄ gases is given as the standard error of the mean of each sequence, multiplied by t, Student's factor for a 95 % confidence limit.

More recently, a variability between different LSVEC aliquots at the IAEA premises has been found which has been attributed to interaction with atmospheric CO₂ and water, leading to a small increase in $\delta^{13}\text{C}_{\text{VPDB}}$ (S. Assonov, pers. comm.). The final $\delta^{13}\text{C}_{\text{VPDB}}$ value assigned to the specimen used in our laboratory might therefore be subject to a (small) adjustment, which will directly affect the calibrated methane values. We have tentatively made allowance for the effect by including an estimated uncertainty of ± 0.03 ‰ for the LSVEC specimen used in the BGC ISOLAB experiments (Table 2). A final adjustment can only be made when a suitable solution to this issue has been found, agreed upon and published.

Similar to the $\delta^{13}\text{C}_{\text{VPDB}}$ calibration, the methane hydrogen isotopes have been calibrated using the international reference water samples VSMOW2 and SLAP2 with $\delta^2\text{H}_{\text{VSMOW}}$ of 0.0 and -427.5 ‰, respectively. Both of these reference materials have small uncertainties of ± 0.3 ‰ for $\delta^2\text{H}_{\text{VSMOW}}$ (IAEA, (2009), Brand et al., (2014)). The uncertainties are included in the error values for the Master-CH₄ gases Megan and Merlin, which are otherwise obtained in an analogous fashion as described above: CH₄ and H₂O sample injections are made within the same sequence. The observed water $\delta^2\text{H}$ values are then used to fix the $\delta^2\text{H}$ -CH₄ results to the VSMOW/SLAP scale. Again, the final $\delta^2\text{H}$ -CH₄ values for the Master-CH₄ gases are presented as the standard error of the sequence means, multiplied by t, Student's factor for a 95 % confidence limit.

The secondary or "calibration" CH₄ gases have been directly analysed against the Master-CH₄ gases using identical procedures and equipment, switching the 10-port valve depicted in Fig. 2. The final errors evaluated for the gas-to-gas comparisons are obtained by the standard error of the calculated mean delta values, multiplied again by t, Student's factor for a 95 % confidence interval. The uncertainty of the respective master methane gas is included in the combined error using standard error propagation. Possible scale contraction effects were treated routinely by measuring against a pair of reference water samples and applying a scale adjustment. This effect is part of the measurement uncertainty.



2.5 Producing synthetic isotope reference standards from pure CH₄ and CH₄-free air

The ISOLAB at the Max-Planck-Institute for Biogeochemistry operates a system to dilute pure CO₂ with CO₂-free air to atmospheric CO₂ mole fraction without isotopic fractionation (Ghosh et al., 2005). This system is referred to as ARAMIS and is used in this study to prepare atmospheric reference gases for CH₄. We dilute an aliquot of calibrated CH₄ with CH₄-free air to atmospheric CH₄ mole fractions (~2 ppm) in 5-L glass flasks with a final filling pressure of 1.8 bar absolute. The CH₄-free matrix air has been target-mixed from ultra-pure constituents. It contains N₂, O₂, N₂O and Kr at atmospheric mixing ratios. Krypton was added to this matrix air to account for the measurement artefact during GC-IRMS analysis of CH₄ for $\delta^{13}\text{C}$ (Schmitt et al., 2013), so that the produced isotope reference gases reflect a natural air sample as closely as possible. A sensitive analysis using high-precision gas-chromatography yielded an upper limit for the CH₄ mixing ratio of 0.5 ppb. Further details on the matrix air and the krypton added are provided in Table 1. The $\delta^2\text{H}$ and $\delta^{13}\text{C}$ isotope ratios of CH₄ in the reference gases produced in the dilution experiments are measured with an average standard deviation (1σ) of 1.2 ‰ and 0.14 ‰, respectively, which is similar to the precision of the analytical system (1.0 ‰ and 0.12 ‰ for $\delta^2\text{H-CH}_4$ and $\delta^{13}\text{C-CH}_4$, respectively (W.A. Brand, 2015, pers. comm.) and suggests reproducible dilution of CH₄.

2.6 Comparison of isotope scales at MPI-BGI and IMAU

A new system to measure both carbon and hydrogen isotope ratios in atmospheric methane has recently been built at MPI-BGC, (W.A. Brand, pers. comm.). This system is referred to as *i*SAAC, in abbreviation for *i*sotope *S*ystem for *A*nalyzing *A*tmospheric *C*onstituents. *i*SAAC has been operational since 2012 with an external precision of 1.0 ‰ and 0.12 ‰ for $\delta^2\text{H-CH}_4$ and $\delta^{13}\text{C-CH}_4$, respectively, as determined using the performance chart technique (Werner and Brand, 2001) over a period of 3 years. We calibrated our measurements using atmospheric working standards referred to as “Carina-1” and “Carina-2” (Table 1). The “Carina” gases have been calibrated at the Institute for Marine and Atmospheric Research in Utrecht (IMAU), using the analytical setup described by Brass and Röckmann, (2010) and Sapart et al., (2011). The history of the isotope scale at IMAU is described in detail by Brass and Röckmann (2010). We arbitrarily chose “Carina-1” as master reference gas for the *i*SAAC system. While “Carina-1” and “Carina-2” agree well in $\delta^{13}\text{C-CH}_4$, their previously calibrated $\delta^2\text{H-CH}_4$ value differs by ~4 ‰. This offset has been confirmed as an artefact of the calibration at IMAU.

The synthetic isotope reference gases that were produced from previously calibrated CH₄ gases were analysed on *i*SAAC against “Carina-1”. The results of these measurements are compared to the calibration results of the pure CH₄ gases so that the differences between the calibrated isotope ratios and the measurements against “Carina-1” indicate the offset between the new and the previous isotope scales in Jena (Sect. 3.3).

30



3 Results

3.1 Results for Master-CH₄ gas calibrations on the international VSMOW and VPDB isotope scales

Our Master-CH₄ gases “Megan” and “Merlin” have been calibrated for their carbon and hydrogen isotope ratios against international reference material and in house working standards (Table 2). The results are shown in Table 3. “Megan” was used as Master-CH₄ gas for all initial experiments and calibrations until it was accidentally vented to ambient in May 2015. In order to compensate the loss, “Merlin” was lifted in reference hierarchy from calibration CH₄ gas to Master-CH₄ gas as successor of “Megan”. Therefore, Merlin was calibrated using two different methods, once against primary reference materials and once against a Master-CH₄ gas. Therefore, “Merlin” is listed twice in Table 3 with results that agree within the combined uncertainty.

Both “Megan” and “Merlin” are fossil CH₄ gases with typical dual isotope ratios for this source category (e.g. Quay et al., (1999), Mikaloff Fletcher et al., (2004)). The two Master-CH₄ gases are similar in $\delta^2\text{H-CH}_4$ with isotope ratios of $-168.0 \pm 0.6 \text{ ‰}$ and $-165.7 \pm 0.6 \text{ ‰}$, respectively. The calibrated $\delta^{13}\text{C-CH}_4$ isotope ratios of “Megan” and “Merlin” are $-40.75 \pm 0.07 \text{ ‰}$ and $-39.07 \pm 0.07 \text{ ‰}$, respectively.

15

20

25

30



Table 3: Results of CH₄ isotope calibrations. Each CH₄ gas is named in column 1 while its function is defined in column 2. Calibrated δ²H-CH₄ and δ¹³C-CH₄ values including combined uncertainties are listed in column 3 and 4, respectively. Note that results for “Merlin” are listed twice, once from its calibration against primary reference materials and once from calibration against Megan, the 1st calibrated Master-CH₄ gas. “Fossil” and “Biogenic” have previously been calibrated elsewhere (Sperlich et al., 2012). The results of the previous calibration are marked with (*) and are compared to the results of this method as discussed in the main text. Martha-1 and Mike-1 were intermittent gases used in the calibration comparison (Table 4) but were then mixed to be similar or more enriched in δ²H than tropospheric CH₄.

Gas name	Function	δ ² H-CH ₄ [‰]	δ ¹³ C-CH ₄ [‰]
Megan	Master	-168.1 ± 0.6	-40.75 ± 0.07
Merlin	Master	-165.7 ± 0.6	-39.07 ± 0.07
Merlin	Calibration	-164.1 ± 1.2	-39.14 ± 0.19
Martha-1	Calibration	-176.6 ± 1.3	-48.79 ± 0.06
Martha-2	Calibration	+36.2 ± 1.2	-48.87 ± 0.15
Mike-1	Calibration	+44.5 ± 1.1	-40.79 ± 0.08
Mike-2	Calibration	-80.3 ± 0.7	-42.74 ± 0.10
Merida	Calibration	-171.7 ± 1.1	-60.28 ± 0.13
Fossil	Calibration	-171.9 ± 1.6	-39.72 ± 0.18
Fossil	Comparison	-170.1 ± 0.7 (*)	-39.56 ± 0.04 (*)
Melly	Calibration	-177.5 ± 1.4	-69.87 ± 0.11
Minion	Calibration	-182.7 ± 1.2	-58.06 ± 0.11
Merkur	Calibration	-195.8 ± 1.2	-43.03 ± 0.10
Biogenic	Calibration	-319.8 ± 1.4	-56.55 ± 0.11
Biogenic	Comparison	-317.4 ± 0.7 (*)	-56.37 ± 0.04 (*)

10

3.2 Results for CH₄ gas calibrations against Master-CH₄ gases

The calibration results of all secondary CH₄ gases are shown in Table 3. Our gases include gases from natural and commercial gas sources as well as isotopically spiked CH₄ mixtures. The gases cover large ranges in carbon (-70 to -39 ‰) and hydrogen isotope ratios (-320 to +36 ‰), which include the isotopic composition of tropospheric CH₄. For δ²H-CH₄ this was achieved by spiking fossil CH₄ gases with pure CH₃D to yield “Martha-2” and “Mike-1”. In order to create a CH₄ gas with tropospheric δ²H-CH₄, “Mike-1” was diluted with a fossil CH₄ gas to produce “Mike-2”, thereby using up “Mike-1”.



3.3 Results for isotopic measurements of CH₄ in synthetic air standards

Aliquots of the previously calibrated CH₄ gases were diluted with CH₄-free air (Sect. 2.4) for analysis on a setup named “iSAAC” that has recently been developed at MPI-BGC for the analysis of dual isotope ratios of CH₄ in air samples (Sect. 2.5). In iSAAC, the diluted CH₄ reference gases are treated as unknown samples and referenced against “Carina-1”. The difference between the calibration results of the pure CH₄ gases and the diluted CH₄ gases is calculated by $\delta_{iSAAC} - \delta_{pure}$, which shows the calibration difference between the two different isotope reference materials. Table 4 displays the observed differences between the two calibration methods.

Our experiments show good agreement between the $\delta^{13}C$ -CH₄ calibrations, but a systematic offset in the calibrations of δ^2H -CH₄. While the average offset in $\delta^{13}C$ -CH₄ of $+0.02 \pm 0.08$ ‰ is smaller than the measurement uncertainty, the average offset in δ^2H -CH₄ accounts for $+4.0 \pm 1.1$ ‰, which is significant.

The measurements of “Biogenic” for $\delta^{13}C$ -CH₄ showed a larger discrepancy between the two methods. The cause of this discrepancy could not be resolved. It may be due to an unknown impurity in the bulk CH₄ gas that only impacts on the analysis of the pure CH₄ gas as iSAAC separates CH₄ chromatographically. A sudden drift in the “Biogenic” cylinder could be another reason for this discrepancy, however, this is not reflected in the analysis of “Biogenic” as pure CH₄ gas over time. A pressure drop in the cylinder that is independent of CH₄ consumption during analysis is not detected. Because the discrepancy exceeds the measurement uncertainty by a factor of 4, we consider this result as an outlier and refrained from using it in further calculations.

Table 4: Comparison of the previous isotope scale for CH₄ in air of the iSAAC system as defined by “Carina-1”, which was calibrated by IMAU (Brass and Röckmann) and the new scale as based on the CH₄ calibrations presented in this study. The first column shows the name of the pure CH₄ gas that has been diluted with CH₄-free air for analysis as atmospheric sample (see Table 3). Columns 2 and 3 show the difference $\delta_{iSAAC} - \delta_{pure}$ for δ^2H -CH₄ and $\delta^{13}C$ -CH₄, respectively. The bottom line shows the mean difference with 1σ uncertainty. The cause for the larger value in $\delta^{13}C$ -CH₄ of “Biogenic” is unknown. Therefore, this value is marked with (°) and is not considered for the determination of the scale difference.

Gas name	δ^2H -CH ₄ [‰]	$\delta^{13}C$ -CH ₄ [‰]
Megan	3.8	0.00
Merlin	4.0	-0.01
Minion	2.7	-0.03
Melly	4.3	0.13
Mike-1	5.7	-0.06
Martha-1	3.2	-0.05
Fossil	5.1	0.14
Biogenic	3.0	0.40 (°)
Average	$+4.0 \pm 1.1$	$+0.02 \pm 0.08$



4 Discussion

We have used standard on-line IRMS techniques to calibrate pure CH₄ for $\delta^2\text{H}$ and $\delta^{13}\text{C}$ on international isotope reference scales. The pure methane gases were injected into the same isotope analysis system as water and carbonate reference materials and, thus, subjected to the same analytical conditions.

- 5 The oxidation of CH₄ to CO₂ (and H₂O) is usually accepted to produce no artefact. However, chemically CH₄ is rather stable; it requires high temperatures and a surplus of oxygen in order to drive the combustion reaction to completion. A major complication arises when yields are lower than 100 % (Merritt et al., 1995). This would allow for unreacted CH₄, which is a potent source of protonation in the IRMS ion source (Anicich, 1993). With CO₂ present, this results in the formation of CO₂H⁺, an important isobaric interference on the m/z 45 mass position, where $\delta^{13}\text{C}$ is measured. This source of analytical error is
- 10 important especially for analytical systems that don't separate CO₂ and CH₄ after the combustion on a chromatographic column or by cryogenic means. In the MPI-BGC systems, we use post-combustion chromatographic column to separate N₂ and CO₂ peaks, which also separates any residual methane from CO₂. In addition, we carefully checked for traces of methane left over from the reaction by monitoring the m/z 15 ion current carefully for un-converted CH₄ and found the combustion to be quantitative.
- 15 The introduction of carbonates to quantitatively release the CO₂ at high temperatures has been demonstrated to yield high precision results (Coplen et al., 2006). A considerable advantage is that the oxygen isotopic composition is altered completely. By “roasting”, it is forced to be very similar for all samples introduced into the combustion furnace, thus making the ¹⁷O correction identical as well. Hence, any ambiguity arising from the necessary correction for extracting $\delta^{13}\text{C}$ values from the m/z 45 ion current tends to cancel. The technique has been used to revise the VPDB scale by adding LSVEC as a second
- 20 scaling point (Coplen et al., 2006).

For hydrogen, we chose an analogue approach to process H₂O and CH₄ through the high-temperature reactor of the TC/EA-IRMS system. This technique has been investigated earlier in detail and shown to result in quantitative conversion of H₂O and methane at reactor temperatures above 1350 °C (Gehre et al., 2004). Possible artefacts can arise mainly from the different surface activities of H₂O versus CH₄ before the conversion to H₂ (and CO or carbon) occurs. For water, this can lead to memory

25 effects, which needs to be taken into account quantitatively either by correcting for it or by discarding initial injections (Werner and Brand, 2001). We found a further, minor dependence of $\delta^2\text{H}$ -H₂O to the septum temperature. In the appendix we describe experimental details to ensure optimal conditions for making the H₂O and CH₄ high temperature reactions directly compatible for quantitative $\delta^2\text{H}$ assessment and calibration.

We achieved the results presented in Table 3 for “Fossil” and “Biogenic” with a large number of analyses. Both gases had

30 already been calibrated by one of the authors in an earlier study using the more classical approach, i.e. combusting pure methane quantitatively off line and sampling the resulting CO₂ and H₂O for consecutive isotope analysis (Sperlich 2012). In this study, the CH₄ derived CO₂ was analysed by dual inlet IRMS while the H₂O was analysed by either TC/EA-IRMS or cavity-ring-down spectroscopy. The results of both calibration methods agree in the order of the combined uncertainty of both



methods for $\delta^2\text{H-CH}_4$ and $\delta^{13}\text{C-CH}_4$ (Table 3). The values from our new calibration for “Biogenic” and “Fossil” appear to be slightly more depleted in $\delta^2\text{H-CH}_4$ (average difference of 2.1 ‰) and in $\delta^{13}\text{C-CH}_4$ (average difference of 0.17 ‰), suggesting that the discrepancy could be systematic. We conclude from the broad agreement with the previously published results on “Fossil” and “Biogenic” that our new method is capable to calibrate CH_4 gases accurately to the international isotope scales.

5 Our new results are based on a large number of individual CH_4 conversions and measurements, using GC-IRMS methods that are used for routine analysis since more than a decade. In comparison, the previously published values of Sperlich et al., (2012) were derived from only four analyses per gas, converted in an off-line reactor. We suggest that the method presented here is more accurate and robust compared to the method of Sperlich et al., (2012).

The calibration of “Merlin” against “Megan” as Master- CH_4 as well as against primary reference standards provides another
10 indication of the accuracy of our analytical methods. The difference between the two calibrations of “Merlin” accounts for 1.6 ‰ in $\delta^2\text{H-CH}_4$ and for 0.07 ‰ in $\delta^{13}\text{C-CH}_4$, which is of the order of the analytical uncertainty. We interpret the good agreement as indicator that the presented methods for CH_4 , H_2O and carbonate analysis do not create any significant isotope fractionation and suggest that our CH_4 gases are accurately calibrated. Note that this conclusion is a best estimation scenario that cannot be tested further without CH_4 reference material.

15 The total propagated uncertainty of the isotope calibration is smaller than or similar to the uncertainty of most analytical systems to measure CH_4 isotope ratios in air or ice core samples and can therefore help to increase the compatibility between international laboratories.

20 5 Conclusions

The number of laboratories that measure isotope ratios of atmospheric CH_4 is growing and combining data from multiple laboratories could lead to more powerful interpretative approaches of the combined data sets. However, the analysis of data from multiple laboratories is currently hampered by the lack of reference materials that enable the community to produce a unified data set.

25 To solve this problem, we accurately referenced twelve pure CH_4 gases to the international isotope scales for carbon and hydrogen and diluted aliquots of eight calibrated CH_4 gases with CH_4 -free air in 5-L glass flasks. These synthetic gas mixtures were tested for their use as standards for CH_4 isotope ratios by comparing separately established isotope scales. From the combined results, the inter-laboratory calibration offsets of the two scales could be established reliably.

Our synthetic atmospheric reference gases for the isotopic composition of CH_4 will be made available to the atmospheric
30 monitoring community upon request. Using these reference gases in multiple laboratories will help unifying the data sets of the atmospheric monitoring community, thus enabling compatible isotope ratio data sets of atmospheric methane on a global scale over time.



Acknowledgements

Financial support was provided within the European Commission projects InGOS and IMECC. N. A. M. Uitslag's visit in Jena was made available by the Erasmus program. Support by Huilin Chen and Harro Meijer (U-Groningen) is gratefully appreciated. We are indebted to Ingeborg Levin for making archived air samples from Neumayer station available for comparison measurements and to Gordon Brailsford for valuable comments during the preparation of the manuscript.

Appendix A: Preliminary experiments for calibration of CH₄ with H₂O for δ²H

The injection of H₂O samples into the reactor is critical because it is prone to isotope fractionation (Werner and Brand, 2001). This isotope fractionation is caused by system memory due to adhesion of injected H₂O to the reactor walls. The isotope fractionation can be overcome by repetitive injections of H₂O samples with identical isotopic composition, thereby overwriting the memory effect until it reaches a marginal level. For H₂O analyses under constant analytical conditions (e.g. constant reactor temperature), the adhesion effect is a function mainly of the amount of injected H₂O sample. However, the effect on the isotopic composition furthermore scales with the isotopic difference between two consecutive samples (Gehre et al., 2004). Because there is no adhesion of the sample during CH₄ analysis, this memory effect is specific only to the analysis of H₂O in our study. Therefore, system memory of H₂O could propagate into the CH₄ calibrations. Memory effects are identified in a series of replicate H₂O measurements and are corrected for by modelling the memory function as described in Gehre et al., (2004) and Brand et al., (2009a) on a routine basis, as our system has been used for isotope analysis of H₂O samples for more than a decade. We conclude that our results are free of artefacts from memory effects.

Isotope fractionation during the analysis of the reference waters can also be caused by insufficiently heated septa (Gehre et al., 2004). We injected 106 H₂O samples while we increased the septum temperatures in 9 steps from 76°C to 137°C. In general, we observed a δ²H enrichment with increasing Septum temperature, however, there seems to be a second effect in the course of increasing temperature. A systematic variation is apparent at temperatures above 90°C with stabilising δ²H-H₂O values at septum temperatures around 130°C (Fig. A1, blue circles). At the three highest temperatures, δ²H-H₂O averages at -62.6 ± 0.5 %. The stabilising δ²H-H₂O at high temperatures suggests quantitative H₂O processing without significant isotope fractionation, in line with (Gehre et al., 2004). Note the disagreement with the three δ²H-H₂O values at the lower temperature range (red diamonds) that show a matching slope but an offset that does not align with the δ²H-H₂O pattern observed at temperatures above 90°C. We cannot explain this offset but can only speculate that the first experiments with a heated septum had a stronger initial impact on the measured isotopic composition. Note that the total septum temperature effect is in the order of twice the typical measurement uncertainty and is thus difficult to observe.





Figure A1: The $\delta^2\text{H}$ variation of H_2O injections with septum temperatures. Blue circles show average values that fall onto a polynomial fit while red diamonds indicate the three values at low septum temperatures that include the offset. The error bars show 1σ standard deviations and the grey-dashed lines indicate the typical precision limit of 1 ‰ for $\delta^2\text{H-H}_2\text{O}$ analysis (Gehre et al., 2004) around the $\delta^2\text{H-H}_2\text{O}$ of -62.45 ‰ as estimated by the polynomial fit for the operating septum temperature of 130°C.

10 Quantitative conversion of both CH_4 and H_2O in the high temperature reactor is of utmost importance for our study, because incomplete conversion causes isotopic fractionation in the reaction products (e.g. Burgoyne and Hayes (1998), Hilkert et al., (1999) Gehre et al., (2004)). The temperature of the reactor is critical for the efficiency of the conversion process. We performed an experiment with CH_4 and H_2O injections at different reaction temperatures and show the results in Fig. A2. For water injections we observe a pronounced, non-linear $\delta^2\text{H-H}_2\text{O}$ change of ~ 15 ‰ with reactor temperature increase from 15 1300°C to 1450°C, reaching a plateau above 1400°C. This pattern is consistent with previous observations in both trend and magnitude (Gehre et al., 2004). In contrast, the linear fit for $\delta^2\text{H-CH}_4$ increases by only about 1 ‰ over the 150 K temperature range. However, the slope is statistically insignificant as shown by the 95 % confidence interval of the linear fit (Fig. A2). This analyte-specific isotope variation is also reflected in the area of the H_2O and CH_4 derived H_2 peaks (Fig. A2). While the H_2O -derived H_2 peak areas increase with increasing reactor temperature, the CH_4 -derived H_2 peak areas remain constant within the 20 error bars throughout the experiments. For an unknown reason, three out of six H_2 peaks that resulted from H_2O injections at 1400°C were by 10-15 standard deviations smaller than the remaining three peaks. We present the averages and 1σ standard of the H_2 peaks with and without removal of these outliers in Fig. A2, which shows the exceptional pattern at 1400°C. Despite of this peak size variability, the isotopic composition of all H_2O injections at 1400°C is in good agreement. Our experiments indicate that high reactor temperatures of 1450°C are required especially for quantitative conversion of H_2O , while the effects 25 of reactor temperature on both yield and the isotopic composition of CH_4 -derived H_2 are comparably small. Therefore, we



operate the reactor at a temperature of 1450°C to guarantee quantitative conversion without isotope fractionation of both, H₂O (Gehre et al., 2004) and CH₄ (Burgoyne and Hayes, (1998), Hilkert et al., (1999)).



5 Figure A2: The variation in isotopic composition and peak area of H₂ derived from H₂O and CH₄ injections at reactor
 temperatures between 1300-1450°C. Top and bottom panels show experiments for H₂O and CH₄, respectively. Isotope ratios
 are shown in blue for H₂O and green for CH₄ and refer to the left hand axes. Average H₂ peak areas are indicated by grey
 crosses and refer to the right hand axes. All error bars indicate the 1σ standard deviation. The red diamond shows the average
 peak area and respective 1σ standard deviation when the three outliers are included (see Appendix text). Y-axes ranges are
 10 matched between top and bottom panels to enable direct comparison of the temperature effect for H₂O and CH₄. Equations
 describe the fits in both panels, displayed by dashed lines. The continuous lines in the bottom panel indicate the 95 %
 confidence interval of the linear fit.

Appendix B: Experiments for calibration of CH₄ with CaCO₃ and Li₂CO₃

15 Incomplete CH₄ combustion results in measurement artefacts, because CH₄-derived CO₂ and unconverted CH₄ enter the IRMS
 simultaneously, enabling the formation of CO₂H⁺ fragments that contribute to the signal on the *m/z* 45 detector (Werner and



Brand, 2001). In order to test the completeness of CH₄ combustion, we set the IRMS to monitor *m/z* 15 during the analysis of a CH₄ sample. We estimate that <0.1 % of an injected CH₄ sample reaches the IRMS unconverted, suggesting quantitative combustion of CH₄.

The analysis of carbonates can suffer from a well characterised blank contribution that is due to the carbon impurities within the tin capsules that are used for the aliquotation of the carbonates (Werner et al., 1999). In contrast, no such blank is expected when samples are analysed without tin capsules, as would be the case for gaseous CH₄ samples. In order to fulfil the principle of identical treatment between analyses of carbonate reference materials and CH₄ samples, we added empty tin capsules to each CH₄ analysis and applied the routine blank correction to all measurements.

To test for the completeness of carbonate digestion, we added small amounts of tungsten trioxide (WO₃) to the carbonate samples during weighing (about 1:1 by weight) to increase the instantaneous reaction temperature and to provide additional oxygen during the liberation of CO₂ from different carbonates. The addition of WO₃ did improve the peak shape for the analysis of BaCO₃ (Table 2) but had no effect on the results of our isotope ratio measurements. Other carbonates (CaCO₃ and Li₂CO₃) did not suffer from a broadened peak shape. This suggests that the carbonate digestion is not limited by either temperature or oxygen availability. Therefore, we conclude that the digestion of carbonate samples is quantitative and we refrained from adding WO₃ during the CH₄ calibration measurements.

References

- Anicich, V. G.: Evaluated bimolecular ion-molecule gas phase kinetics of positive ions for use in modeling planetary atmospheres, cometary comae, and interstellar clouds, *J. Phys. Chem. Ref. Data*, 22, 1469-1569, 1993.
- Bergamaschi, P., Schupp, M., and Harris, G. W.: High-precision direct measurements of ¹³CH₄/¹²CH₄ and ¹²CH₃D/¹²CH₄ ratios in atmospheric methane sources by means of a long-path tunable diode-laser absorption spectrometer, *Appl. Optics*, 33, 7704-7716, 1994.
- Bock, M., Schmitt, J., Beck, J., Schneider, R., and Fischer, H.: Improving accuracy and precision of ice core δD(CH₄) analyses using methane pre-pyrolysis and hydrogen post-pyrolysis trapping and subsequent chromatographic separation, *Atmos. Meas. Tech.*, 7, 1999-2012, 2014.
- Brand, W. A., Coplen, T. B., Aerts-Bijma, A. T., Bohlke, J. K., Gehre, M., Geilmann, H., Groning, M., Jansen, H. G., Meijer, H. A. J., Mroczkowski, S. J., Qi, H. P., Soergel, K., Stuart-Williams, H., Weise, S. M., and Werner, R. A.: Comprehensive inter-laboratory calibration of reference materials for δ¹⁸O versus VSMOW using various on-line high-temperature conversion techniques, *Rapid Communications in Mass Spectrometry*, 23, 999-1019, 2009a.
- Brand, W. A., Coplen, T. B., Vogl, J., Rosner, M., and Prohaska, T.: Assessment of international reference materials for isotope-ratio analysis (IUPAC Technical Report), *Pure Appl. Chem.*, 86, 425-467, 2014.
- Brand, W. A., Huang, L., Mukai, H., Chivulescu, A., Richter, J. M., and Rothe, M.: How well do we know VPDB? Variability of δ¹³C and δ¹⁸O in CO₂ generated from NBS19-calcite, *Rapid Commun. Mass Spectrom.*, 23, 915-926, 2009b.
- Brass, M. and Röckmann, T.: Continuous-flow isotope ratio mass spectrometry method for carbon and hydrogen isotope measurements on atmospheric methane, *Atmos. Meas. Tech.*, 3, 1707-1721, 2010.



- Brooks, P. D., Geilmann, H., Werner, R. A., and Brand, W. A.: Improved precision of coupled $\delta^{13}\text{C}$ and $\delta^{15}\text{N}$ measurements from single samples using an elemental analyser/isotope ratio mass spectrometer combination with a post-column six-port valve and selective CO_2 trapping; improved halide robustness of the combustion reactor using CeO_2 , *Rapid Communications in Mass Spectrometry*, 17, 1924-1926, 2003.
- 5 Burgoyne, T. W. and Hayes, J. M.: Quantitative production of H_2 by pyrolysis of gas chromatographic effluents, *Analytical Chemistry*, 70, 5136-5141, 1998.
- Coplen, T. B., Brand, W. A., Gehre, M., Gröning, M., Meijer, H. A. J., Toman, B., and Verkouteren, R. M.: New guidelines for delta C-13 measurements, *Anal. Chem.*, 78, 2439-2441, 2006.
- Craig, H.: The Geochemistry of the Stable Carbon Isotopes, *Geochim. Cosmochim. Acta*, 3, 53-92, 1953.
- 10 Dumke, I., Faber, E., and Poggenburg, J.: Determination of Stable Carbon and Hydrogen Isotopes of Light Hydrocarbons, *Analytical Chemistry*, 61, 2149-2154, 1989.
- Friedman, I.: Deuterium content of natural waters and other substances, *Geochimica Et Cosmochimica Acta*, 4, 89-103, 1953.
- Eyer, S., Tuzson, B., Popa, M. E., van der Veen, C., Röckmann, T., Rothe, M., Brand, W. A., Fisher, R., Lowry, D., Nisbet, E. G., Brennwald, M. S., Harris, E., Zellweger, C., Emmenegger, L., Fischer, H., and Mohn, J.: Real-time analysis of $\delta^{13}\text{C}$ - and $\delta\text{D-CH}_4$ in ambient air with laser spectroscopy: method development and first intercomparison results, *Atmos. Meas. Tech. Discuss.*, 8, 8925-8970, doi:10.5194/amt-d-8-8925-2015, 2015
- 15 Gehre, M., Geilmann, H., Richter, J., Werner, R. A., and Brand, W. A.: Continuous flow $^2\text{H}/^1\text{H}$ and $^{18}\text{O}/^{16}\text{O}$ analysis of water samples with dual inlet precision, *Rapid Communications in Mass Spectrometry*, 18, 2650-2660, 2004.
- 20 Ghosh, P., Patecki, M., Rothe, M., and Brand, W. A.: Calcite- CO_2 mixed into CO_2 -free air: a new CO_2 -in-air stable isotope reference material for the VPDB scale, *Rapid Commun. Mass Spectrom.*, 19, 1097-1119, 2005.
- Gröning, M., Duren van, M., and Andreescu, L.: Metrological characteristics of the conventional measurement scales for hydrogen and oxygen stable isotope amount ratios: the δ -scales. Proceedings of an International
- 25 Workshop on "Combining and reporting analytical results: The role of traceability and uncertainty for comparing analytical results". Rome, 6-8 March 2006, 2007, 62-72.
- Hilkert, A. W., Douthitt, C. B., Schlüter, H. J., and Brand, W. A.: Isotope ratio monitoring gas chromatography/mass spectrometry of D/H by high temperature conversion isotope ratio mass spectrometry, *Rapid Communications in Mass Spectrometry*, 13, 1226-1230, 1999.
- 30 IAEA: Reference sheet for international reference standards, VSMOW2 and SLAP2, http://nucleus.iaea.org/rpst/Documents/VSMOW2_SLAP2.pdf, 2009.
- Kai, F. M., Tyler, S. C., Randerson, J. T., and Blake, D. R.: Reduced methane growth rate explained by decreased Northern Hemisphere microbial sources, *Nature*, 476, 194-197, 2011.
- Levin, I., Bergamaschi, P., Dorr, H., and Trapp, D.: Stable Isotopic Signature of Methane from Major Sources in
- 35 Germany, *Chemosphere*, 26, 161-177, 1993.
- Levin, I., Veidt, C., Vaughn, B. H., Brailsford, G., Bromley, T., Heinz, R., Lowe, D., Miller, J. B., Poss, C., and White, J. W. C.: No inter-hemispheric $\delta^{13}\text{C}\text{CH}_4$ trend observed, *Nature*, 486, E3-E4, 2012.
- Lowe, D. C., Brenninkmeijer, C. A. M., Brailsford, G. W., Lassey, K. R., Gomez, A. J., and Nisbet, E. G.: Concentration and ^{13}C records of atmospheric methane in New Zealand and Antarctica - Evidence for changes in
- 40 methane sources, *Journal of Geophysical Research-Atmospheres*, 99, 16913-16925, 1994.
- Merritt, D. A., Freeman, K. H., Ricci, M. P., Studley, S. A., and Hayes, J. M.: Performance and optimization of a combustion interface for isotope ratio monitoring gas-chromatography mass-spectrometry, *Analytical Chemistry*, 67, 2461-2473, 1995.
- Mikaloff Fletcher, S. E., Tans, P. P., Bruhwiler, L. M., Miller, J. B., and Heimann, M.: CH_4 sources estimated from atmospheric observations of CH_4 and its $^{13}\text{C}/^{12}\text{C}$ isotopic ratios: 1. Inverse modeling of source processes, *Global Biogeochemical Cycles*, 18, 17, 2004.
- 45



- Quay, P., Stutsman, J., Wilbur, D., Snover, A., Dlugokencky, E., and Brown, T.: The isotopic composition of atmospheric methane, *Global Biogeochemical Cycles*, 13, 445-461, 1999.
- Rella, C. W., Hoffnagle, J., He, Y., and Tajima, S.: Local- and regional-scale measurements of CH₄, δ¹³CH₄, and C₂H₆ in the Uintah Basin using a mobile stable isotope analyzer, *Atmos. Meas. Tech.*, 8, 4539-4559, 2015.
- 5 Renpenning, J., Kummel, S., Hitzfeld, K. L., Schimmelmann, A., and Gehre, M.: Compound-Specific Hydrogen Isotope Analysis of Heteroatom-Bearing Compounds via Gas Chromatography-Chromium-Based High-Temperature Conversion (Cr/HTC)-Isotope Ratio Mass Spectrometry, *Analytical Chemistry*, 87, 9443-9450, 2015.
- Sapart, C. J., Monteil, G., Prokopiou, M., van de Wal, R. S. W., Kaplan, J. O., Sperlich, P., Krumhardt, K. M.,
 10 van der Veen, C., Houweling, S., Krol, M. C., Blunier, T., Sowers, T., Martinerie, P., Witrant, E., Dahl-Jensen, D., and Röckmann, T.: Natural and anthropogenic variations in methane sources during the past two millennia, *Nature*, 490, 85-88, 2012.
- Sapart, C. J., van der Veen, C., Vigano, I., Brass, M., van de Wal, R. S. W., Bock, M., Fischer, H., Sowers, T.,
 Buizert, C., Sperlich, P., Blunier, T., Behrens, M., Schmitt, J., Seth, B., and Röckmann, T.: Simultaneous stable
 15 isotope analysis of methane and nitrous oxide on ice core samples, *Atmos. Meas. Tech.*, 4, 2607-2618, 2011.
- Schiegl, W. E. and Vogel, J. C.: Deuterium Content of Organic Matter, *Earth and Planetary Science Letters*, 7, 307-&, 1970.
- Schmitt, J., Seth, B., Bock, M., Fischer, H., Möller, L., Sapart, C. J., Prokopiou, M., Van der Veen, C.,
 Röckmann, T., and Sowers, T.: On the interference of Kr during carbon isotope analysis of atmospheric methane
 20 using continuous flow combustion - isotope ratio mass spectrometry, *Atmos. Meas. Tech.*, 6, 1425-1445, 2013.
- Schmitt, J., Seth, B., Bock, M., and Fischer, H.: Online technique for isotope and mixing ratios of CH₄, N₂O, Xe
 and mixing ratios of organic trace gases on a single ice core sample, *Atmos. Meas. Tech.*, 7, 2645-2665, 2014.
- Sessions, A. L., Burgoyne, T. W., and Hayes, J. M.: Determination of the H₃ factor in hydrogen isotope ratio
 monitoring mass spectrometry, *Analytical Chemistry*, 73, 200-207, 2001.
- 25 Sperlich, P., Buizert, C., Jenk, T. M., Sapart, C. J., Prokopiou, M., Röckmann, T., and Blunier, T.: An automated
 GC-C-GC-IRMS setup to measure paleoatmospheric δ¹³C-CH₄, δ¹⁵N-N₂O and δ¹⁸O-N₂O in one ice core sample,
Atmos. Meas. Tech., 6, 2027-2041, 2013.
- Sperlich, P., Guillevic, M., Buizert, C., Jenk, T. M., Sapart, C. J., Schaefer, H., Popp, T. J., and Blunier, T.: A
 combustion setup to precisely reference δ¹³C and δ²H isotope ratios of pure CH₄ to produce isotope reference
 30 gases of δ¹³C-CH₄ in synthetic air, *Atmos. Meas. Tech.*, 5, 2227-2236, 2012.
- Sperlich, P., Schaefer, H., Mikaloff Fletcher, S. E., Guillevic, M., Lassey, K. R., Sapart, C. J., Röckmann, T., and
 Blunier, T.: Carbon isotope ratios suggest no additional methane from boreal wetlands during the rapid
 Greenland Interstadial 21.2, *Global Biogeochemical Cycles*, 29, 1962-1976, 2015.
- Stevens, C. M. and Rust, F. E.: The Carbon Isotopic Composition of Atmospheric Methane, *Journal of*
 35 *Geophysical Research-Oceans and Atmospheres*, 87, 4879-4882, 1982.
- Tokida, T., Nakajima, Y., Hayashi, K., Usui, Y., Katayanagi, N., Kajiura, M., Nakamura, H., and Hasegawa, T.:
 Fully automated, high-throughput instrumentation for measuring the δ¹³C value of methane and application of the
 instrumentation to rice paddy samples, *Rapid Communications in Mass Spectrometry*, 28, 2315-2324, 2014.
- Wendeberg, M., Richter, J. M., Rothe, M., and Brand, W. A.: Jena Reference Air Set (JRAS): a multi-point scale
 40 anchor for isotope measurements of CO₂ in air, *Atmos. Meas. Tech.*, 6, 817-822, 2013.
- Werner, R. A. and Brand, W. A.: Referencing strategies and techniques in stable isotope ratio analysis, *Rapid
 Communications in Mass Spectrometry*, 15, 501-519, 2001.
- Werner, R. A., Bruch, B. A., and Brand, W. A.: ConFlo III - An interface for high precision δ¹³C and δ¹⁵N
 analysis with an extended dynamic range, *Rapid Communications in Mass Spectrometry*, 13, 1237-1241, 1999.



- WMO: Report of the 16th WMO/IAEA Meeting on Carbon Dioxide, Other Greenhouse Gases, and Related Measurement Techniques (GGMT-2011), edited by: Brailsford, G., GAW Report 206, Wellington, New Zealand, 2012.
- WMO: Report of the 17th WMO/IAEA Meeting on Carbon Dioxide, Other Greenhouse Gases, and Related Tracers Measurement Techniques (GGMT-2013), edited by: Tans, P. and Zellweger, C., GAW Report 213, Beijing, China, 2014.
- 5