


Article

Monitoring the Wobbe Index of Natural Gas Using Fiber-Enhanced Raman Spectroscopy

Vincenz Sandfort ¹, Barbara M. Trabold ², Amir Abdolvand ^{2,3}, Carsten Bolwien ⁴, Philip St. J. Russell ², Jürgen Wöllenstein ^{1,4} and Stefan Palzer ^{5,*} 

¹ Laboratory for Gas Sensors, Department of Microsystems Engineering–IMTEK, University of Freiburg, Georges-Köhler-Allee 102, 79110 Freiburg, Germany; vincenz.sandfort@imtek.uni-freiburg.de

² Max Planck Institute for the Science of Light, Staudtstraße 2, 91058 Erlangen, Germany;

barbara.trabold@mpl.mpg.de (B.M.T.); aabdolvand@ntu.edu.sg (A.A.); philip.russell@mpl.mpg.de (P.S.J.R.)

³ School of Electrical and Electronic Engineering, Nanyang Technological University, 50 Nanyang Ave, Singapore 639798, Singapore

⁴ Fraunhofer Institute for Physical Measurement Techniques IPM, Heidenhofstraße 8, 79110 Freiburg, Germany; Carsten.Bolwien@ipm.fraunhofer.de (C.B.); Juergen.Woellenstein@ipm.fraunhofer.de (J.W.)

⁵ Department of Computer Science, Universidad Autónoma de Madrid, Francisco Tomás y Valiente 11, 28049 Madrid, Spain

* Correspondence: stefan.palzer@uam.es

Received: 26 October 2017; Accepted: 23 November 2017; Published: 24 November 2017

Abstract: The fast and reliable analysis of the natural gas composition requires the simultaneous quantification of numerous gaseous components. To this end, fiber-enhanced Raman spectroscopy is a powerful tool to detect most components in a single measurement using a single laser source. However, practical issues such as detection limit, gas exchange time and background Raman signals from the fiber material still pose obstacles to utilizing the scheme in real-world settings. This paper compares the performance of two types of hollow-core photonic crystal fiber (PCF), namely photonic bandgap PCF and kagomé-style PCF, and assesses their potential for online determination of the Wobbe index. In contrast to bandgap PCF, kagomé-PCF allows for reliable detection of Raman-scattered photons even below 1200 cm^{-1} , which in turn enables fast and comprehensive assessment of the natural gas quality of arbitrary mixtures.

Keywords: Raman spectroscopy; wobbe index; photonic crystal fiber; kagomé; fiber enhanced; natural gas

1. Introduction

The analysis of complex gas matrices at reasonable cost is an often required but seldom accomplished task and applications for such a technology range from air quality management to process monitoring as well as safety and security applications. One of the world's biggest markets for multi-gas analysis equipment is in the natural gas sector, where fast and reliable determination of the Wobbe index [1–4] is much sought after. The Wobbe index is used as a measure for the interchangeability of fuel gases and indicates their combustion energy.

Currently, Fourier transform infrared spectroscopy (FTIR [5–8]), gas chromatography (GC [9,10]) and mass spectrometry (in particular coupled with GC: GC-MS [11–13]) are the most common laboratory-based tools whenever the constituents of the gas mixture are unknown but their identification is required. Other, more cost-effective technologies, such as e.g., metal oxide based gas sensors [14] or standalone flame ionization detectors [15], show a high degree of sensitivity to many gases, but do not allow specific detection and identification of individual gas components. Among the optical methods, infrared absorption spectroscopy tools such as tunable diode laser spectroscopy

(TDLAS) or non-dispersive infrared spectroscopy (NDIR) allow selective and sensitive detection of individual components [16–20]. However, they are unsuitable in terms of total cost in scenarios with many different gases because many laser sources are needed to perform TDLAS on the distinct absorption lines or many spectral filters in case of NDIR.

Raman spectroscopy offers the opportunity to simultaneously measure gas mixtures [21–24] using a single light source. It is based on the Raman effect, experimentally first observed by Krishnan and Raman in 1928 [25], and relies on the change of energy of photons that are scattered during the interaction with matter. While it is well known and widespread for analyzing solids in crystallography and liquids in chemistry [26], it is less often used in commercial products for gas analysis as compared to absorption spectroscopy techniques. However, important applications include the analysis of reacting flows [25] or atmospheric temperature determination [27–30]. This is due to the typically small Raman scattering cross-sections $\sigma_{Raman}(\lambda_0) \sim (\frac{1}{\lambda_0})^4$ (λ_0 is the excitation wavelength), which are several orders of magnitude below the absorption cross-section σ_{Abs} of infrared active lines used in TDLAS [31]. Compared to solids or liquids the Raman scattering rate in gas is typically three orders of magnitude lower due to the lower density. According to [32–34] the detected Raman intensity I_{Raman} depends on the excitation laser intensity I_{Laser} , the number of molecules N in the illuminated volume V and the effective interaction length l_{eff} :

$$I_{Raman} = \varepsilon \cdot I_{Laser} \cdot \frac{N}{V} \cdot l_{eff} \cdot \sigma_{Raman} \quad (1)$$

for a Raman apparatus with total detection efficiency ε . The obvious possibilities to increase I_{Raman} are increasing the particle number N , increasing the excitation intensity I_{Laser} , or using shorter excitation wavelengths. For this reason several strategies are being pursued for enhancing the scattered light intensity when probing gases. Enhancing the signal may be achieved using non-linear techniques such as coherent anti-Stokes Raman spectroscopy (CARS [35–37]) or stimulated Raman spectroscopy (SRS [37–39]). While these techniques require at least two laser sources whose frequency difference may be tuned, the Raman signal may be amplified by several orders of magnitude as compared to spontaneous Raman scattering. However, the more complicated optical setups and limitations imposed by the frequency range that may be tuned limit the range of applications outside the laboratory. On the other hand, non-coherent Raman spectroscopy methods do require only one laser source, albeit at a lower scattering intensity. Commercial products for probing gas samples are on the market [40,41] and several strategies are being investigated for enhancing the scattered light intensity when probing gases using a single laser source to improve signal strength and widen the range of possible applications. To this end, multipass cells [42–47] and capillary fibers [48,49] allow for amplifying the Raman signals. One prominent possibility relies on using optical resonators to enhance the light intensity available for scattering (CERS-cavity enhanced Raman spectroscopy) [50,51]. Another approach uses hollow-core photonic crystal fiber (HC-PCF) to enhance the number of probed molecules by increasing the interaction length with the gas and additionally guiding the scattered photons (FERS-fiber enhanced Raman spectroscopy) [21,22,52,53]. In this way a photonic bandgap (PBG) HC-PCF has been used to increase the Raman scattering signal by several orders of magnitude [54] depending on the type and length of the fiber. To estimate the enhancement factor of a HC-PCF as compared to a focused laser beam one may compare the effective interaction lengths l_{eff} . In case of a free-space beam the collecting optics can only deliver a certain part of the scattered signal to the spectrometer, limiting l_{eff} to roughly twice the Rayleigh length. For a free running laser beam at 532 nm with a beam diameter of 4 μm and a Rayleigh range z_R of 23.6 μm the effective interaction length is thus $l_{eff} = 2 \cdot z_R = 2 \cdot \pi W_0^2 / \lambda = 47.2 \mu\text{m}$. Comparing this to the interaction length of a HC-PCF 1 m in length yields an amplification factor of approximately 21,000. However, a direct comparison between Raman systems with and without [55] fibers is difficult because the total detection efficiency is dependent on the quality of the Raman light collection optics, which is not standardized.

In the past the enhancement capabilities of FERS have been employed to demonstrate the feasibility of using Raman spectroscopy in demanding applications such as breath gas analysis [52], isotope selective greenhouse gas detection [56], and natural and biogas analysis. Among the various possible applications, one of the biggest markets is the fast and reliable analysis of natural gas compositions [57–59]. While in prior work the use of a cavity-enhanced setup to detect some natural gas components such as methane, propane, ethane, and hydrogen sulfide was reported [51], less effort has been devoted to assessing its suitability for online monitoring of natural gas. However, natural gas supply and transport utilities are keen to determine the so-called Wobbe index [1,3,4]. The Wobbe index is a measure for the heating value of a gas composition and is the most-frequently used indicator for the quality of natural gas and the interchangeability of different gas types [2,60,61]. Determination of the Wobbe index requires identification and quantification of the hydrocarbon components, which typically feature Raman shifts between 800 cm^{-1} and 3100 cm^{-1} , and other further constituents like CO_2 and N_2 . Using Raman spectroscopy may offer a convenient, fast, and reliable alternative overcoming the drawbacks of established technologies like FTIR, TDLAS, and NDIR. However, some important components of natural gas mixtures, e.g., ethane, propane and *n*-butane, feature Raman peaks in the region below 1200 cm^{-1} , which are covered by the background signal from the silica cladding. The glass background signal makes it difficult to record signals from the gas, even though out the light emitted from the cladding or, alternatively, analyzing peaks in the CH-stretching region around 3000 cm^{-1} , may improve the situation. But due to the many CH-stretching modes of many hydrocarbons in this region this requires a very fine resolution of the measurement to resolve the different modes of different gases.

Here, we demonstrate for the first time the use of a kagomé fiber for FERS analysis of natural gas as a much simpler alternative to suppress the glass background, simply because the light fraction in the glass is predicted to be about 10 dB lower than in PBG-PCF. On that basis, we investigate the possibility of using FERS as a fast, reliable and cost efficient method for determining the Wobbe index. The performance of kagomé-PCF and PBG-PCF is compared in terms of acquisition time and signal to noise level when analyzing a natural gas composition. To evaluate the practical issue of measurement frequency when using HC-PCF, the gas exchange times are determined for both fiber types and an experimental approach to minimize the effects of dead volumes is presented. Other techniques to reduce the evacuation times of photonic crystal fibers include micro-channel machining [62,63] might be used additionally. While kagomé-PCF is known for a low Raman signal background [64], the typically larger core diameter results in a weaker signal strength. The fiber length is optimized, which means making a compromise between short gas exchange times and high Raman signal levels.

2. Setup and Methods

A system for Raman spectroscopy [54] has been adapted for gas measurements in HC-PCF. Confinement of gas and guidance of light are achieved in hollow-core PBG-PCF and kagomé-PCF, respectively. Both are expected to exhibit different behavior in terms of gas exchange rates, background signals and enhancement factors. The PBG-PCF used is a “HC-580” fiber from NKT Photonics (Birkerød, Denmark), with a core radius of $2.25\text{ }\mu\text{m}$. The alternatively available “HC-532” has a lower attenuation at 532 nm, but for Raman measurements up to 3500 cm^{-1} the fiber has to deliver a little attenuation of the Stoke photons with wavelengths up to 632 nm. This is why the “HC-580” is more suitable for this specific task. Its performance is compared to a custom-made kagomé-PCF [64,65] with a core radius of $12\text{ }\mu\text{m}$. Varying fiber lengths between 5 cm and 80 cm are used. Figure 1 shows scanning electron micrographs (SEMs) of the in-coupling facet for both fiber types.

The experimental setup is depicted in Figure 2a. A frequency doubled Nd: YAG laser (COMPASS 215M50) at 532 nm is used as light source and the light is coupled into the HC-PCFs using a Raman edge filter (Semrock SEM-LP03-532RE-25) and a microscope objective. The input fiber facet is also used as output port for the backward scattered Raman photons. In this way, the objective collimates the Raman-scattered light and the edge filter disposes of the excitation wavelength. To enable simultaneous

coupling of light and gas into the fiber, it is mounted in a gas-tight adapter (Figure 2b) that can withstand up to 14 bar. The dead volume of 200 μL may be rinsed through a purging valve, which opens up the dead volume to air. Optical access is gained through a 0.5 mm thick sapphire window. Two connectors attached to the adapter provide gas-tight interfaces to the optical fiber and a gas delivery system, respectively. The latter is derived from a more sophisticated setup for control of gas compositions in laboratory environments [66]. The fiber is connected via an SMA905-thread and the gas delivery system via an M3-thread for gas connectors. The other fiber end facet is left open to a gas extraction system operating just below ambient pressure.




Figure 1. Scanning electron microscopy images of the kagomé-PCF (left) and a photonic bandgap PCF (right). The core diameters are 12 μm and 2.25 μm , respectively. Please note the 5-fold difference in magnification.




Figure 2. (a) Schematic drawing of the setup. The excitation light is depicted in green and the Raman-scattered in yellow; (b) Plumbing diagram for the gas delivering system with the hollow fiber (yellow); (c) Quantum efficiency of the Andor iDus401-BVF camera. The red-shaded region marks the spectral range used.

The back-scattered light is then directed onto a spectrometer, with a 50- μm -pinhole filtering out parts of the stray light. A diffraction grating with 1800 lines/mm is used to spatially separate the Raman-scattered photons according to their wavelengths. The spectrum is imaged onto a CCD camera (iDus401-BVF, Andor Technology, Belfast, UK), which records the intensity of scattered light around the pump beam frequency $\bar{\nu}_0 \approx 18,800 \text{ cm}^{-1}$ in the range from $\bar{\nu}_0 = -100 \text{ cm}^{-1}$ to $\bar{\nu}_0 = +3500 \text{ cm}^{-1}$ using the full length of the CCD array of (1024×127) pixels with a pixel size of $(26 \times 26) \mu\text{m}^2$ at a distance of 107.5 mm between grating and camera resulting in a spectral resolution of $\sim 8 \text{ cm}^{-1}$. The camera is set to full vertical binning mode, 16.25 μs vertical pixel shift time, 50 kHz horizontal pixel readout rate and $1.0\times$ pre-amp gain. With these camera settings, every electron is converted to 4.7 A/D-counts on average. For every pixel, the quantum efficiency $QE(\lambda)$ of the camera for the corresponding wavelength is given in Figure 2. The detected Raman scattering rate R_{Raman} provided in the graphs below is calculated using the expression:

$$R_{Raman} = \frac{N_{ph}}{t_{Int}} = \frac{N_e}{QE(\lambda) \cdot t_{Int}} = \frac{C}{4.7 \cdot QE(\lambda) \cdot t_{Int}} \quad (2)$$

where N_{ph} is the number of incoming photons, N_e is the number of excited electrons, t_{Int} is the integration time and C is the number of A/D counts. The total scattering rate for each gas species is determined by integrating the values of all corresponding pixels of the respective Raman peak. Coupling into the fiber was achieved by a 3-dimensional microscope stage. All measurements are performed at an ambient temperature of 23 $^{\circ}\text{C}$ and an outlet pressure of ~ 1 bar at the exhaust. The natural gas composition used in the experiments is a sample blend purchased from Air Liquide (Paris, France) and is composed of the most prominent, typically occurring natural gas constituents according to the list in Table 1. To change the gas composition inside the fiber the gas mixture is fed in via the gas connection at an absolute pressure of up to 4 bar. To evaluate and optimize the gas exchange rate caused by the dead volume, experiments with and without using the purging valve were performed.

Table 1. Composition of the gas mixture used in the experiments according to the manufacturer certificate.

Gas	Specified Molar Concentration [mol %]	Measurement Uncertainty [mol %]
<i>n</i> -Butane	0.985	0.020
Propane	3.374	0.067
Ethane	9.50	0.190
Methane	83.93	0.321
Nitrogen	0.4044	0.0081
Carbon dioxide	1.806	0.036

With the setup in Figure 2, the full Raman spectrum can be analyzed in a single measurement, enabling determination of the gas composition. This makes the technique suitable for on-line determination of the Wobbe index. In general, inferior and superior limits are given for the Wobbe index, namely the inferior Wobbe index W_i and the superior Wobbe index W_s . They can be calculated from the lower heating value H_i and the gross heating value H_s as [67]:

$$W_i = \frac{H_i}{\sqrt{\rho/\rho_0}} \quad (3a)$$

$$W_s = \frac{H_s}{\sqrt{\rho/\rho_0}} \quad (3b)$$

where ρ is the gas density and ρ_0 the density of dry air. The lower and gross heating values for the contained gases, as well as their individual densities for 100% purity, are listed in Table 2. The Wobbe index of a gas mixture is the concentration-weighted mean value of the individual Wobbe indices.

Here, the relative concentrations of the individual components are deduced from the measured Raman spectra, which allows real-time calculation of the Wobbe index.

Table 2. List of lower and gross heating values as well as the gaseous density of the natural gas components [67–70].

Gas	Lower Heating Value H_i [MJ/m ³]	Gross Heating Value H_s [MJ/m ³]	Density ρ [kg/m ³]
<i>n</i> -Butane	122.910	133.119	2.71
Propane	93.215	101.242	2.01
Ethane	64.345	70.293	1.36
Methane	35.883	39.819	0.72
Nitrogen	0	0	1.25
Carbon dioxide	0	0	1.98

3. Results

Filling both fiber types with the natural gas mixture at an input pressure of 2 bar and a length of 80 cm results in the spectra shown in Figure 3. The light coupled into each fiber was held constant for all measurements at 27 mW for the PBG-PCF and 28 mW for the kagomé-PCF, respectively. Notably, the PBG-PCF generates a strong background signal, ranging from approximately 0 to 1700 cm⁻¹, even partly saturating the camera at an integration time of 0.5 s. This is due to the non-negligible portion of light propagating in the SiO₂ cladding material [71–73], which, as a solid, features orders of magnitude higher Raman gain as compared to a gas. In the kagomé-PCF, however, the glass background is strongly suppressed because the electric field penetrating the glass is typically around 10 dB lower as compared to PBG-PCF [74].

The corresponding molecular spectra obtained using a kagomé-PCF show a much reduced scattering rate, i.e., the integrated nitrogen peak is about 4 times lower than in the PBG-PCF. This is due to the larger core diameter of the kagomé-PCF, which results in a smaller acceptance angle for the scattered light to be guided in the core. Nevertheless, the reduced background scattering from glass allows the detection of natural gas components including butane and propane below 1700 cm⁻¹, which could not be detected using the PBG fiber.

In order to evaluate the practical suitability and attainable measurement speed, the gas exchange times and the signal strength for different lengths and input pressures are measured. Therefore, pure nitrogen gas is substituted by the natural gas mixture and the Raman peak of N₂ recorded during the exchange. A sample measurement for 2 bar input pressure with closed purging valve is shown in Figure 4a. The nitrogen Raman peak is localized around 2331 cm⁻¹, which results in a peak extending from 2311 cm⁻¹ to 2355 cm⁻¹ detected on the camera, i.e., 11 pixels when using the configuration of the spectrometer setup above. To evaluate the total N₂ scattering rate the background signal of the camera is subtracted before integrating over all involved pixels, to obtain the total Raman scattering rate displayed in Figure 4b. The t_{90} time, i.e., the it takes to reach 90% of the new steady-state reading, is deduced by applying an exponential decay fit (red dashed curve), resulting in 47 s for the measurement in Figure 4b.




Figure 3. Spectra of the Raman scattering of natural gas for an input laser power of 27 mW. Displayed are scattered photons per second over the Raman shift in wavenumbers, i.e., inverse centimeters. Note the different vertical scales for the upper and middle panel. The large glass background for the PBG fiber can be seen from approximately 100 cm^{-1} to 1700 cm^{-1} . The lower panel displays the position of the different Raman peaks of the gas components. The relative Raman scattering cross-section for each molecule is normalized to that of nitrogen [37,75–77].




Figure 4. Procedure for evaluating the total nitrogen Raman scattering rate, illustrated using a 0.8 m length of kagomé-PCF at an absolute input pressure of 2 bar. (a) Nitrogen peak around 2331 cm^{-1} at various times during the gas exchange; (b) Transient plot of the integrated Raman scattering rate during gas exchange, which starts at $t = 0\text{ s}$. An exponential fit yields the t_{90} time.

To determine the gas exchange times for different fiber lengths and input pressures, Raman spectra are recorded at a frequency of 1 Hz for the kagomé-PCF and 2 Hz for the PBG-PCF. The resulting t_{90} times are shown in Figure 5. There is a negligible correlation between the input pressure and the gas exchange time and also the fiber length does not appear to have a major influence. When entering

the fiber, the gas velocity increases due to the conservation of mass principle and at the same time the Venturi effect lowers the pressure behind the fiber. At a fixed input pressure the mass flow will not increase beyond a limiting condition, i.e., the so-called critical valve or choked flow effect [78], which ultimately results in a constant flow velocity inside the fiber. The small fiber core diameters and the applied pressure differences of $\Delta p \geq 1$ bar at a minimum pressure of 1 bar lead to a Knudsen number $K_n \ll 1$ and laminar flow [79], in contrast to free molecular flow that appears for Knudsen numbers $K_n \gg 1$. For photonic crystal fibers operated in varying regimes these flow regions are discussed in detail in [80,81].

The mean gas exchange time here is 49.1 s with a standard deviation of 16.3 s for the kagomé-PCF and 1466 s with a standard deviation of 332 s for the PBG-PCF. That means gas flow into the kagomé-PCF is about 30 times faster than into the PBG-PCF. As the kagomé-PCF's core area is 28 times larger, the diameter of the fiber core is identified as the decisive factor, which is in line with expectations from the critical valve effect. The fibers used in here feature core holes much larger than the surrounding cladding holes, which is why the gas exchange will predominantly happen where the gas is detected as well. While we do not anticipate any interference by slower gas exchange rates in the cladding, should such effects occur they may be eliminated by selectively closing the cladding holes of the input facet for gas [82].




Figure 5. Gas exchange times of (a) kagomé-PCF and (b) PBG-PCF at lengths of 0.1 m, 0.25 m, 0.5 m and 0.8 m. Please note the different scales on the y-axis.

Because the gas in the dead volume has to be exchanged via the HC-PCF, the measured gas exchange times are on the order of minutes, which makes the design of small dead volumes a priority in this type of setup. Another method consists of replacing the gas in the dead volume with the help of the purging valve such that only the gas volume actually filling the fiber needs to flow through the fiber. To this end, experiments have been performed employing PBG-PCF with lengths of 15 cm and 27 cm. The experimental protocol is altered so as to allow purging of the dead volume for five seconds prior to starting data capture of the Raman signal. Because the purge valve has a diameter of 1 mm the dead volume is purged mostly through the purging valve, while the gas composition inside the hollow fiber remains constant.

In Figure 6 the time-dependent Raman signal for nitrogen is shown after the purge valve has been closed. The nitrogen molecules inside the fiber are now replaced by the natural gas mixture within seconds. Comparing the temporal behavior in the Figures 4 and 6 reveals important differences between the two approaches. Gas exchange via the dead volume (cf. Figure 4) results in an exponential decay associated with the mixing process of both gas samples in the dead volume. If the dead volume is rinsed first, the decrease in scattering rate is linear with time. Employing a linear fit to the N_2 Raman scattering rate in Figure 6, the slope b can be determined and used to calculate the gas exchange time $t_{90} = -0.9/b$ shown in Figure 7a. From this, the gas flow velocity v_{ge} is determined and plotted in Figure 7b. The exchange time now linearly depends on the pressure and is reduced from 1466 s on average to tens of seconds, i.e., by more than two orders of magnitude for the PBG-PCF.




Figure 6. Normalized nitrogen scattering rate during exchange with natural gas from PBG-PCF with lengths of (a) 27 cm and (b) 15 cm. The purge valve is used to quickly change the gas in the dead volume in front of the fiber. Once it is closed, the measurement starts. Notably, the decrease in scattering rate is now linear during the gas exchange process.




Figure 7. (a) Gas exchange time of the PBG fiber as a function of the input pressure when employing the purging valve; (b) Gas exchange velocity for various pressures and fiber lengths of 15 cm and 27 cm, respectively.

The results in Figure 7 correspond well with basic considerations where the natural gas pushes the nitrogen out of the fiber without mixing. Assuming a laminar flow inside the hollow fiber as well as the validity of Hagen-Poiseuille's equation one would expect a gas flow velocity according to [79]:

$$v_m(\Delta P) = \frac{1}{4\eta} \cdot \frac{R^2}{l} \cdot \Delta P = G_v \cdot \Delta P, \quad (4)$$

where R is the fiber core radius, l the fiber length, η the viscosity of the gas filling, i.e., $\eta_{N_2} = 17.7 \mu\text{Pa s}$ for N_2 at room temperature, and $\Delta P = P_i - P_o$ the pressure difference between input P_i and output pressure P_o . The velocity increase per bar is labeled G_v .

G_v can be determined experimentally from the slope of the linear fits in Figure 7b. However, while this yields the correct order of magnitude for the gas exchange time, the fiber length appears to have no bearing on the gas velocity inside the fiber. This is also due to the choked valve effect, which limits the mass flow. In fact, applying a linear fit to the gas exchange velocity as a function of the input pressure results in slopes for both lengths that are equal within the margin of error. Table 3 summarizes these findings. Using Equation (4) is then suitably only to give an estimation of the correct order of magnitude for G_v . This model is also applicable for the kagomé-PCF, which should result in even faster gas exchanged times (c.f. Equation (4)).

Table 3. Theoretical and measured pressure-dependent gas velocities.

Fiber Length/cm	$G_v/m\ s^{-1}\ bar^{-1}$	$G_{v,theo}/m\ s^{-1}\ bar^{-1}$
15	$(8.0 \pm 1.0) \times 10^{-3}$	12.05×10^{-3}
27	$(8.6 \pm 0.4) \times 10^{-3}$	6.69×10^{-3}

From Figure 7a and Equation (4), it is clear that a short gas exchange time requires a short fiber length, which conflicts with the need for a long fiber to maximize the Raman signal. In the following, the influence of the fiber length on the detected scattering rate is investigated so as to determine the ideal fiber length. The light coupled into each fiber length was held constant at 27 mW for the PBG-PCF and 28 mW for the kagomé-PCF. The measured Raman scattering signals are shown in Figure 8. The nonlinear dependence of the Raman scattering rate on fiber length is mainly caused by fiber loss, resulting in two separate effects. Firstly, the exponential decrease in pump intensity along the fiber reduces the number of generated Raman photons. Secondly, the Raman scattered photons are attenuated on the way back to the output port. These losses are further increased by coupling of Raman scattered photons into higher order guided modes that have higher losses than the fundamental mode. Pump and Stokes attenuation counteract the linear increase of the scattered photon number with the fiber length (see Equation (1)). For this reason the scattering rate saturates with increasing fiber length.




Figure 8. Integrated nitrogen signal plotted over the fiber length at 2 bar pressure for (a) kagomé-PCF and (b) PBG-PCF. A logarithmic fit is applied to take account of the two-way-attenuation in the fiber. Insets show the loss of the fundamental mode specified by the manufacturer.

Based on the theoretical considerations above, the data can be fitted with the following model (red curve in Figure 8):

$$R_{Raman}(L) \sim \int_0^L J_0 \cdot e^{-(\alpha_P + \alpha_S) \cdot l} dl = \frac{J_0}{\alpha_P + \alpha_S} (1 - e^{-(\alpha_P + \alpha_S) \cdot L}), \quad (5)$$

where L is the fiber length, α_P and α_S the pump and Stokes attenuation coefficients, respectively, and $J_0 = I_0/m$ is initial laser power per meter of the fiber. The resulting fit parameters are summarized in Table 4. The value for $\alpha_S + \alpha_P$ in Table 4 is considerably higher than predicted by the loss curves in the insets of Figure 8. This happens because higher-order guided modes losses range from few dB/m to above 100 dB/m loss [83].

Table 4. Fit parameters for both types of fibers as well as the offset value for 0 m fiber length.

Fiber Type	$J_0/(\alpha_S + \alpha_P)/Hz$	$\alpha_S + \alpha_P/m^{-1}$
kagomé	5535 ± 764	4.41 ± 0.69
PBG	$29,587 \pm 5693$	4.94 ± 1.25

To quantify the performance of both fibers as a function of their lengths, the differential Raman scattering rate gain $R'(l) = dR_{Raman}(l)/dl$ has been calculated, i.e., the slope of the curves in Figure 8. The ideal fiber length l_{ideal} has been defined as $R'(l_{ideal}) = 0.5 R'(0)$ and is indicated by the green marks in Figure 8. This leads to ideal FERS fiber lengths of 14.1 cm for the PBG-PCF and 15.6 cm for the kagomé-PCF.

Using the Raman scattering rates from Figure 8 for the ideal fiber length, the limit of detection (LOD) for both fiber types has been determined for the most prominent components of natural gas. The LOD is defined as [84]:

$$LOD = 3 \cdot \frac{c}{SNR}, \quad (6)$$

where c is the gas concentration and SNR is the measured signal-to-noise ratio. Noise is mainly generated by the standard deviation of the signal and the background noise of the camera caused by stray light and random detection events. The signal-to-noise ratio is calculated from the signal amplitude S and the square root of the the standard deviation of the error σ [85]: $SNR = \frac{S}{\sqrt{\sigma}}$. Since the photon scattering rate is an integral over a multiple of camera pixels, the derived background noise is the sum of the used pixels for every band. The resulting excitation-intensity dependent LODs are listed in Table 5 in ppm at 2 bar pressure and 1 s integration time, together with their associated Raman shifts in wavenumbers and the relative Raman scattering cross sections at that shift. The Raman bands used for the calculation of the hydrocarbon values are in the region from 800 cm^{-1} to 1600 cm^{-1} , due to the mentioned band overlaps in the C-H stretching region around 3000 cm^{-1} . With a high resolution spectrometer and chemometrics the upper bands may be used as well albeit at the cost of using considerable more complex data analysis methods.

Table 5. Limit of detection in ppm for optimized fiber lengths and 80 cm long fibers and 1 s integration time at 2 bar input pressure. Higher laser power leads to a linear decrease in the LOD.

Gas	PBG-PCF LOD @ 27 mW Pump Power [ppm]		Kagomé-PCF LOD @ 28 mW Pump Power [ppm]		Wave-Number [cm ⁻¹]	Relative Raman Scattering Intensity
	14.1 cm Length	80 cm Length	15.6 cm Length	80 cm Length		
<i>n</i> -Butane	-	-	4140	730	827	1.9 [75]
Propane	-	-	5950	1050	870	1.6 [37,75]
Ethane	2230	440	6050	1070	993	1.2 [37,75]
Methane	62,900	12,300	75,200	13,300	1535	0.1 [75]
Nitrogen	570	110	1350	240	2331	1 [76,77]
Carbon Dioxide	1630	320	4300	760	1388	1.1 [37,75]

Increasing the input pressure leads to a linear increase of the number of molecules inside any given fiber. The pressure-dependent scattering rate of nitrogen is plotted in Figure 9 for both fiber types at a length of 0.8 m. Both show a linear increase with the number of molecules and according to Equation (1) also with the laser power. For example, the LOD for nitrogen at the currently used laser power and an input pressure of 2 bar is 1350 ppm, using a kagomé-PCF with 15.6 cm length. This value can, however, be much improved by increasing the pressure, the fiber length or the laser power. At 14 bar input pressure, 15.6 cm kagomé-PCF length and 1 W pump power inside the fiber would yield an LOD of approximately 10 ppm for nitrogen using the same setup and an integration time of 1 s.

To demonstrate the potential of this technique the Wobbe index of the natural gas mixture is determined using the relative concentrations of all gas matrix components. Since this is the decisive factor, higher pressures can be used for more accurate readings using the same setup. Using the findings from this paper, the kagomé-PCF is more suitable for this application than the PBG-PCF because of the easier access to spectral region below 1700 cm^{-1} . Besides the longer gas exchange time, PBG-PCFs generate a strong glass background, which does not allow facile detection of, e.g., *n*-butane and propane. As these contribute significantly to the heating value of natural gas, accurate determination of the Wobbe index is easier possible in the kagomé-PCF.




Figure 9. Integrated nitrogen signal plotted over the pressure for (a) kagomé-PCF and (b) PBG-PCF. The red curve shows a linear regression.

Figure 10 shows the Wobbe index obtained from the spectral data taken at a frequency of 1 Hz in a 80 cm long kagomé-PCF at 28 mW laser power and 2 bar input pressure during gas exchange. The purging valve was closed during this measurement, so that about 1 min is needed to reach the equilibrium value in order to highlight the capability of an online determination of the Wobbe index. The speed of Wobbe index determination in a real-world setting is limited by the integration time, the venting, subsequent gas exchange time, and the computing time to determine the intensity of each peak of the Raman spectrum. In conclusion, a Wobbe index determination once every 30 s should be feasible using the currently employed setup and a kagomé-PCF. Assuming that all gas components may be clearly identified a calibration of the reading for the Wobbe index is achieved via the relative strength of individual gas peaks. The measured intensity is therefore set to the known concentration. For comparison, the theoretical Wobbe index is calculated from the gas mixture specifications (horizontal line in Figure 10). The measurement converges well to the theoretical value, which indicates good accuracy of our measurement.




Figure 10. Inferior and superior Wobbe indices during gas exchange in the 80 cm long kagomé-PCF with 2 bar pressure and 28 mW laser power. Dots mark the indices calculated from concentration measurements. The curves are exponential growth fits. The horizontal lines mark the Wobbe indices calculated from the manufacturer specifications for the chosen natural gas mixture. Please note that the venting valve was not employed in this measurement.

4. Conclusions

In this paper we have demonstrated the use of hollow-core PCF to perform fiber enhanced Raman spectroscopy (FERS) in order to enable fast and reliable determination of the Wobbe index of natural gas. To evaluate the suitability in real-world deployments, the influence of fiber length, gas exchange time and input pressure have been determined for two prominent fiber types, namely hollow-core photonic bandgap PCF and kagomé-PCF. While the detected Raman scattering rate of the PBG-PCF is about four times larger, the background signal of the fiber material prevents the easy detection of gaseous molecules with Raman shifts below 1700 cm^{-1} , such as *n*-butane and propane. The region around 3000 cm^{-1} could in principle be used to this end but the strong overlap of many hydrocarbons make the signal analysis challenging and prone to inaccuracies. While higher resolution spectrometers for light analysis as well as chemometrics methods may be deployed to resolve this issue, this would incur considerably higher computational and calibration costs. This is the main reason why kagomé-PCF is a better choice for determining the Wobbe index because both *n*-butane and propane feature well-separated rovibrational bands that may be singled out easily in the region below 1700 cm^{-1} . Using fibers with even larger core diameter would allow even faster exchange rates. In particular, single-ring fibers [65] could be used, which have very similar properties to kagomé-PCF, also for large core diameters of up to $80\text{ }\mu\text{m}$. Based on an ideal fiber length, the detection limit of the most prominent gas components has been determined for both types of fiber. The gas exchange times were optimized by purging the dead volume in front of the fiber, allowing exchange of the gas filling of the fiber in a few seconds. The results may pave the way for a robust and fast determination of the natural gas quality at low overall cost. Laser power and gas pressure can be adapted on the basis of the LODs given in this manuscript, so as to design custom systems based on the required detection limit. The optomechanical design of this FERS system may be adapted to higher pressures and gas temperatures such that the maximum ratings are governed by the optical fibers, which can withstand pressures up to 100 bar and temperatures of about $600\text{ }^{\circ}\text{C}$. The techniques employed here make an on-line Wobbe index measurement with a Raman gas system feasible and results are in good agreement with theoretical values. Furthermore, since the kagomé fibers feature low background scattering other gases such as carbon monoxide (CO), oxygen (O₂), and hydrogen (H₂), may be detected with the current setup. For gases with a Raman shifts exceeding 3500 cm^{-1} , such as water (H₂O) at 3652 cm^{-1} , the spectrometer has to be recalibrated. However, water adsorption on the inner glass surface of the fiber has been identified as a possible source for increased optical losses in hollow fibers, which is why long term stable operation should prevent the exposure of the fiber to water molecules.

Supplementary Materials: The following are available online at <http://www.mdpi.com/1424-8220/17/12/2714/s1>, Figure S1: Full spectra of pure nitrogen for the kagomé-type and PBG fiber. Fiber length is 80 cm, applied pressure 2 bar and laser power coupled into the fiber 28 mW and 27 mW respectively, Table S1: Wavenumbers and scattering intensity relative to nitrogen of important natural gas components.

Acknowledgments: This work was supported by the Georg H. Endress foundation. The authors wish to thank Arno Feist for performing the SEM characterization. S.P. acknowledges funding from the Community of Madrid under grant number 2016-T1/AMB-1695. The article processing charge was funded by the German Research Foundation (DFG) and the University of Freiburg in the funding programme Open Access Publishing.

Author Contributions: The experiment has been designed by V.S., C.B. and S.P.; V.S. has performed the measurements. A.A. produced the kagomé fiber. V.S. and S.P. analyzed the data. V.S., B.M.T. and S.P. wrote and J.W., C.B., P.S.J.R. edited the manuscript.

Conflicts of Interest: The authors declare no competing financial interest.

References

1. Florisson, O.; Burrie, P.H. Rapid determination of the Wobbe index of natural gas. *J. Phys. E* **1989**, *22*, 123–128. [[CrossRef](#)]
2. Loubar, K.; Rahmouni, C.; Le Corre, O.; Tazerout, M. A combustionless determination method for combustion properties of natural gases. *Fuel* **2007**, *86*, 2535–2544. [[CrossRef](#)]

3. Zachariah-Wolff, J.L.; Egyedi, T.M.; Hemmes, K. From natural gas to hydrogen via the Wobbe index: The role of standardized gateways in sustainable infrastructure transitions. *Int. J. Hydrogen Energy* **2007**, *32*, 1235–1245. [[CrossRef](#)]
4. Klimstra, J. Interchangeability of Gaseous Fuels—The Importance of the Wobbe-Index. *Exp. Fluids* **1986**, *95*, 962–972.
5. Griffiths, P.R.; de Haseth, J.A. *Fourier Transform Infrared Spectrometry*; John Wiley & Sons, Inc.: Hoboken, NJ, USA, 2007; ISBN 9780470106310.
6. Gremlich, H.-U. Infrared and Raman Spectroscopy. In *Handbook of Analytical Techniques*; Wiley-VCH Verlag GmbH: Weinheim, Germany, 2008; pp. 465–507, ISBN 9783527618323.
7. Demirgian, J.C. Gas chromatography—Fourier transform infrared spectroscopy—Mass spectrometry. A powerful tool for component identification in complex organic mixtures. *TrAC Trends Anal. Chem.* **1987**, *6*, 58–64. [[CrossRef](#)]
8. Hammer, S.; Griffith, D.W.T.; Konrad, G.; Vardag, S.; Caldwell, C.; Levin, I. Assessment of a multi-species in situ FTIR for precise atmospheric greenhouse gas observations. *Atmos. Meas. Tech.* **2013**, *6*, 1153–1170. [[CrossRef](#)]
9. Grob, R.L.; Barry, E.F. *Modern Practice of Gas Chromatography*, 4th ed.; Grob, R.L., Barry, E.F., Eds.; John Wiley & Sons, Inc.: Hoboken, NJ, USA, 2004; ISBN 0471229830.
10. Eiceman, G.A. Instrumentation of Gas Chromatography. In *Encyclopedia of Analytical Chemistry*; John Wiley & Sons, Ltd.: Chichester, UK, 2006; pp. 1–9, ISBN 9780470027318.
11. Sparkman, O.D.; Penton, Z.; Kitson, F.G. *Gas Chromatography and Mass Spectrometry: A Practical Guide*; Academic Press: Waltham, MA, USA, 2011; ISBN 9780080920153.
12. Repa, P.; Tesar, J.; Gronych, T.; Peksa, L.; Wild, J. Analyses of gas composition in vacuum systems by mass spectrometry. *J. Mass Spectrom.* **2002**, *37*, 1287–1291. [[CrossRef](#)] [[PubMed](#)]
13. Matthews, D.E.; Hayes, J.M. Isotope-ratio-monitoring gas chromatography-mass spectrometry. *Anal. Chem.* **1978**, *50*, 1465–1473. [[CrossRef](#)]
14. Wang, C.; Yin, L.; Zhang, L.; Xiang, D.; Gao, R. Metal oxide gas sensors: Sensitivity and influencing factors. *Sensors* **2010**, *10*, 2088–2106. [[CrossRef](#)] [[PubMed](#)]
15. Holm, T. Aspects of the mechanism of the flame ionization detector. *J. Chromatogr. A* **1999**, *842*, 221–227. [[CrossRef](#)]
16. Hodgkinson, J.; Tatam, R.P. Optical gas sensing: A review. *Meas. Sci. Technol.* **2013**, *24*, 012004. [[CrossRef](#)]
17. Harris, S.J.; Weiner, A.M. Detection of atomic oxygen by intracavity spectroscopy. *Opt. Lett.* **1981**, *6*, 142–144. [[CrossRef](#)] [[PubMed](#)]
18. Wartewig, S.; Schorn, C.; Bigler, P. *IR and Raman Spectroscopy*; Wiley-VCH Verlag GmbH & Co., KGaA: Darmstadt, Germany, 2003; ISBN 352730245X.
19. Scholz, L.; Ortiz Perez, A.; Bierer, B.; Eaksen, P.; Wöllenstein, J.; Palzer, S. Miniature low-cost carbon dioxide sensor for mobile devices. *IEEE Sens. J.* **2017**, *17*, 2889–2895. [[CrossRef](#)]
20. Knobelspies, S.; Bierer, B.; Ortiz Perez, A.; Wöllenstein, J.; Kneer, J.; Palzer, S. Low-cost gas sensing system for the reliable and precise measurement of methane, carbon dioxide and hydrogen sulfide in natural gas and biomethane. *Sens. Actuators B Chem.* **2016**, *236*, 885–892. [[CrossRef](#)]
21. Buric, M.P.; Chen, K.P.; Falk, J.; Woodruff, S.D. Improved sensitivity gas detection by spontaneous Raman scattering. *Appl. Opt.* **2009**, *48*, 4424–4429. [[CrossRef](#)] [[PubMed](#)]
22. Hanf, S.; Keiner, R.; Yan, D.; Popp, J.; Frosch, T. Fiber-Enhanced Raman Multigas Spectroscopy: A Versatile Tool for Environmental Gas Sensing and Breath Analysis. *Anal. Chem.* **2014**, *86*, 5278–5285. [[CrossRef](#)] [[PubMed](#)]
23. Westenskow, D.R.; Coleman, D.L. Can the Raman scattering analyzer compete with mass spectrometers: An affirmative reply. *J. Clin. Monit.* **1989**, *5*, 34–36. [[CrossRef](#)] [[PubMed](#)]
24. Wagenen, R.A.; Westenskow, D.R.; Benner, R.E.; Gregonis, D.E.; Coleman, D.L. Dedicated monitoring of anesthetic and respiratory gases By Raman scattering. *J. Clin. Monit.* **1986**, *2*, 215–222. [[CrossRef](#)]
25. Raman, C.V. A New Type of Secondary Radiation. *Nature* **1928**, *121*, 501–502. [[CrossRef](#)]
26. Zoubir, A. (Ed.) *Raman Imaging*; Springer Series in Optical Sciences; Springer: Berlin/Heidelberg, Germany, 2012; Volume 168, ISBN 978-3-642-28251-5.
27. Eckbreth, A.C.; Bonczyk, P.A.; Verdick, J.F. Combustion diagnostics by laser Raman and fluorescence techniques. *Prog. Energy Combust. Sci.* **1979**, *5*, 253–322. [[CrossRef](#)]

28. Lederman, S. The use of laser Raman diagnostics in flow fields and combustion. *Prog. Energy Combust. Sci.* **1977**, *3*, 1–34. [[CrossRef](#)]
29. Behrendt, A.; Reichardt, J. Atmospheric temperature profiling in the presence of clouds with a pure rotational Raman lidar by use of an interference-filter-based polychromator. *Appl. Opt.* **2000**, *39*, 1372. [[CrossRef](#)] [[PubMed](#)]
30. Cooney, J.; Pina, M. Laser radar measurements of atmospheric temperature profiles by use of Raman rotational backscatter. *Appl. Opt.* **1976**, *15*, 602–603. [[CrossRef](#)] [[PubMed](#)]
31. Mooradian, A. Laser Raman Spectroscopy. *Science* **1970**, *169*, 20–25. [[CrossRef](#)] [[PubMed](#)]
32. Demtröder, W. *Laserspektroskopie 2. Experimentelle Techniken*, 6th ed.; Springer Spektrum: Berlin, Germany, 2013; ISBN 9783662442166.
33. Larkin, P. *Infrared and Raman Spectroscopy: Principles and Spectral Interpretation*; Elsevier: San Diego, CA, USA, 2011; ISBN 9780123870186.
34. Long, D.A. *The Raman Effect: A Unified Treatment of the Theory of Raman Scattering by Molecules*; John Wiley & Sons Ltd.: Chichester, UK, 2002; ISBN 0471490288.
35. Regnier, P.R.; Moya, F.; Taran, J.P.E. Gas Concentration Measurement by Coherent Raman Anti-Stokes Scattering. *AIAA J.* **1974**, *12*, 826–831. [[CrossRef](#)]
36. Tolles, W.M.; Nibler, J.W.; McDonald, J.R.; Harvey, A.B. A Review of the Theory and Application of Coherent Anti-Stokes Raman Spectroscopy (CARS). *Appl. Spectrosc.* **1977**, *31*, 253–271. [[CrossRef](#)]
37. Bougeard, D.; Buback, M.; Cao, A.; Gerwert, K.; Heise, H.M.; Hoffmann, G.G.; Jordanov, B.; Kiefer, W.; Korte, E.-H.; Kuzmany, H.; et al. *Infrared and Raman Spectroscopy: Methods and Applications*; Schrader, B., Ed.; Wiley-VCH Verlag GmbH: Weinheim, Germany, 2008; ISBN 3527615423.
38. Matsko, A.B.; Savchenkov, A.A.; Letargat, R.J.; Ilchenko, V.S.; Maleki, L. On cavity modification of stimulated Raman scattering. *J. Opt. B Quantum Semiclass. Opt.* **2003**, *5*, 272–278. [[CrossRef](#)]
39. Bloembergen, N. The Stimulated Raman Effect. *Am. J. Phys.* **1967**, *35*, 989–1023. [[CrossRef](#)]
40. Kaiser Optical Systems AirHead™ Gas-Phase Raman Probe. Available online: http://www.kosi.com/na_en/products/raman-spectroscopy/raman-probes-sampling/airhead-probe-head.php (accessed on 26 October 2017).
41. Kaiser Optical Systems RamanRxn4™ Gas-Phase Analyzer. Available online: http://www.kosi.com/na_en/products/raman-spectroscopy/raman-analyzers/raman-rxn4-gas-phase-analyzer.php (accessed on 26 October 2017).
42. Petrov, D.V. Multipass optical system for a Raman gas spectrometer. *Appl. Opt.* **2016**, *55*, 9521–9525. [[CrossRef](#)] [[PubMed](#)]
43. Hill, R.A.; Mulac, A.J.; Hackett, C.E. Retroreflecting multipass cell for Raman scattering. *Appl. Opt.* **1977**, *16*, 2004–2006. [[CrossRef](#)] [[PubMed](#)]
44. Schlüter, S.; Seeger, T.; Popovska-Leipertz, N.; Leipertz, A. Atemzyklusgenaues Anästhesiegas-Monitoring mit einer laserbasierten Raman-Sonde unter klinischen Bedingungen. *TM Tech. Mess.* **2016**, *83*, 289–299. [[CrossRef](#)]
45. Schlüter, S.; Seeger, T.; Popovska-Leipertz, N.; Leipertz, A. Laserbasierte On-line-Analyse von Biogasen mit einer Raman-Sonde. *TM Tech. Mess.* **2014**, *81*, 546–553. [[CrossRef](#)]
46. Seeger, T.; Schlüter, S.; Lukyanov, G.N. Potentials of Raman based sensor system for an online analysis of human inhale and exhale. *Sci. Tech. J. Inf. Technol. Mech. Opt.* **2015**, *16*, 976–983. [[CrossRef](#)]
47. Schlüter, S.; Krischke, F.; Popovska-Leipertz, N.; Seeger, T.; Breuer, G.; Jeleazcov, C.; Schüttler, J.; Leipertz, A. Demonstration of a signal enhanced fast Raman sensor for multi-species gas analyses at a low pressure range for anesthesia monitoring. *J. Raman Spectrosc.* **2015**, *46*, 708–715. [[CrossRef](#)]
48. Buric, M.P.; Chen, K.; Falk, J.; Velez, R.; Woodruff, S. Raman Sensing of Fuel Gases Using a Reflective Coating Capillary Optical Fiber. *Fiber Opt. Sens. Appl. VI* **2009**, *7316*, 731608-1–731608-8. [[CrossRef](#)]
49. Stone, J. cw Raman fiber amplifier. *Appl. Phys. Lett.* **1975**, *26*, 163–165. [[CrossRef](#)]
50. Neely, G.O.; Nelson, L.Y.; Harvey, A.B. Modification of a Commercial Argon Ion Laser for Enhancement of Gas Phase Raman Scattering. *Appl. Spectrosc.* **1972**, *26*, 553–555. [[CrossRef](#)]
51. Hippler, M. Cavity-Enhanced Raman Spectroscopy of Natural Gas with Optical Feedback cw-Diode Lasers. *Anal. Chem.* **2015**, *87*, 7803–7809. [[CrossRef](#)] [[PubMed](#)]

52. Hanf, S.; Bögözi, T.; Keiner, R.; Frosch, T.; Popp, J. Fast and Highly Sensitive Fiber-Enhanced Raman Spectroscopic Monitoring of Molecular H₂ and CH₄ for Point-of-Care Diagnosis of Malabsorption Disorders in Exhaled Human Breath. *Anal. Chem.* **2015**, *87*, 982–988. [[CrossRef](#)] [[PubMed](#)]
53. Jochum, T.; Rahal, L.; Suckert, R.J.; Popp, J.; Frosch, T. All-in-one: A versatile gas sensor based on fiber enhanced Raman spectroscopy for monitoring postharvest fruit conservation and ripening. *Analyst* **2016**, *141*, 2023–2029. [[CrossRef](#)] [[PubMed](#)]
54. Lambrecht, A.; Bolwien, C.; Herbst, J.; Kühnemann, F.; Sandfort, V.; Wolf, S. Neue Methoden der laserbasierten Gasanalytik. *Chem. Ing. Tech.* **2016**, *88*, 746–755. [[CrossRef](#)]
55. Petrov, D.V.; Matrosov, I.I. Raman Gas Analyzer (RGA): Natural Gas Measurements. *Appl. Spectrosc.* **2016**, *70*, 1770–1776. [[CrossRef](#)] [[PubMed](#)]
56. Yang, X.; Chang, A.S.P.; Chen, B.; Gu, C.; Bond, T.C. High sensitivity gas sensing by Raman spectroscopy in photonic crystal fiber. *Sens. Actuators B Chem.* **2013**, *176*, 64–68. [[CrossRef](#)]
57. Egging, R.G.; Gabriel, S.A. Examining market power in the European natural gas market. *Energy Policy* **2006**, *34*, 2762–2778. [[CrossRef](#)]
58. Kjärstad, J.; Johnsson, F. Prospects of the European gas market. *Energy Policy* **2007**, *35*, 869–888. [[CrossRef](#)]
59. Hulshof, D.; van der Maat, J.-P.; Mulder, M. Market fundamentals, competition and natural-gas prices. *Energy Policy* **2016**, *94*, 480–491. [[CrossRef](#)]
60. Altfeld, K.; Schley, P. Development of natural gas qualities in Europe. *Gas Qual.* **2011**, *152*, 544–550.
61. Abbott, D.J.; Bowers, J.P.; James, S.R. The Impact of Natural Gas Composition Variations on the Operation of Gas Turbines for Power Generation. In Proceedings of the 6th International Conference on the Future of Gas Turbine Technology, Brussels, Belgium, 17–18 October 2012; pp. 1–9.
62. van Brakel, A.; Grivas, C.; Petrovich, M.N.; Richardson, D.J. Micro-channels machined in microstructured optical fibers by femtosecond laser. *Opt. Express* **2007**, *15*, 8731. [[CrossRef](#)] [[PubMed](#)]
63. Hensley, C.; Broaddus, D.H.; Schaffer, C.B.; Gaeta, A.L. Photonic band-gap fiber gas cell fabricated using femtosecond micromachining. *Opt. Express* **2007**, *15*, 6690. [[CrossRef](#)] [[PubMed](#)]
64. Ghenuche, P.; Rammner, S.; Joly, N.Y.; Scharrer, M.; Frosz, M.; Wenger, J.; Russell, P.S.J.; Rigneault, H. Kagome hollow-core photonic crystal fiber probe for Raman spectroscopy. *Opt. Lett.* **2012**, *37*, 4371–4373. [[CrossRef](#)] [[PubMed](#)]
65. Pearce, G.J.; Wiederhecker, G.S.; Poulton, C.G.; Burger, S.; St J Russell, P. Models for guidance in kagome-structured hollow-core photonic crystal fibres. *Opt. Express* **2007**, *15*, 12680–12685. [[CrossRef](#)] [[PubMed](#)]
66. Kneer, J.; Eberhardt, A.; Walden, P.; Ortiz Pérez, A.; Wöllenstein, J.; Palzer, S. Apparatus to characterize gas sensor response under real-world conditions in the lab. *Rev. Sci. Instrum.* **2014**, *85*. [[CrossRef](#)] [[PubMed](#)]
67. ISO/TC 193/SC 1 Analysis of Natural Gas. *ISO 6976:2016: Natural Gas—Calculation of Calorific Values, Density, Relative Density and Wobbe Index from Composition*, 3rd ed.; ISO: Geneva, Switzerland, 2016.
68. Grote, K.-H.; Feldhusen, J. *Dubbel: Taschenbuch für den Maschinenbau*, 23rd ed.; Springer: Berlin/Heidelberg, Germany, 2011; ISBN 9783642173059.
69. Hahne, E. *Technische Thermodynamik: Einführung und Anwendung*; Oldenbourg: München, Germany, 2010; ISBN 9783486592313.
70. Cerbe, G.; Lendt, B.; Brüggemann, K.; Dehli, M.; Gröschl, F. *Grundlagen der Gastechnik Gasbeschaffung—Gasverteilung—Gasverwendung*; Hanser: München, Germany, 2017; ISBN 9783446449657.
71. Hass, M. Raman spectra of vitreous silica, germania and sodium silicate glasses. *J. Phys. Chem. Solids* **1970**, *31*, 415–422. [[CrossRef](#)]
72. McMillan, P., Jr.; Remmele, R. Hydroxyl sites in SiO₂ glass: A note on infrared and Raman spectra. *Am. Mineral.* **1986**, *71*, 772–778.
73. Zotov, N.; Ebbsjö, I.; Timpel, D.; Keppler, H. Calculation of Raman spectra and vibrational properties of silicate glasses: Comparison between Na₂Si₄O₉ and SiO₂ glasses. *Phys. Rev. B* **1999**, *60*, 6383–6397. [[CrossRef](#)]
74. Couny, F.; Benabid, F.; Roberts, P.J.; Light, P.S.; Raymer, M.G. Generation and photonic guidance of multi-octave optical-frequency combs. *Science* **2007**, *318*, 1118–1121. [[CrossRef](#)] [[PubMed](#)]
75. Kiefer, J.; Seeger, T.; Steuer, S.; Schorsch, S.; Weikl, M.C.; Leipertz, A. Design and characterization of a Raman-scattering-based sensor system for temporally resolved gas analysis and its application in a gas turbine power plant. *Meas. Sci. Technol.* **2008**, *19*, 85408. [[CrossRef](#)]

76. Eichmann, S.C.; Weschta, M.; Kiefer, J.; Seeger, T.; Leipertz, A. Characterization of a fast gas analyzer based on Raman scattering for the analysis of synthesis gas. *Rev. Sci. Instrum.* **2010**, *81*, 125104. [[CrossRef](#)] [[PubMed](#)]
77. Fenner, W.R.; Hyatt, H.A.; Kellam, J.M.; Porto, S.P.S. Raman cross-section of some simple gases. *J. Opt. Soc. Am.* **1973**, *63*, 73–77. [[CrossRef](#)]
78. Abbott, M.M. *Perry's Chemical Engineer Handbook*, 8th ed.; Perry, R.H., Green, D.W., Maloney, J.O., Eds.; McGraw-Hill: New York, NY, USA, 2008; ISBN 9780071422949.
79. Bruus, H. *Theoretical Microfluidics*; Oxford University Press: Toronto, ON, Canada, 2008; Volume 18, ISBN 0199235090.
80. Henningsen, J.; Hald, J. Dynamics of gas flow in hollow core photonic bandgap fibers. *Appl. Opt.* **2008**, *47*, 2790. [[CrossRef](#)] [[PubMed](#)]
81. Dicaire, I.; Beugnot, J.-C.; Thévenaz, L. Analytical modeling of the gas-filling dynamics in photonic crystal fibers. *Appl. Opt.* **2010**, *49*, 4604. [[CrossRef](#)] [[PubMed](#)]
82. Zhang, Y.; Shi, C.; Gu, C.; Seballos, L.; Zhang, J.Z. Liquid core photonic crystal fiber sensor based on surface enhanced Raman scattering. *Appl. Phys. Lett.* **2007**, *90*, 2005–2008. [[CrossRef](#)]
83. Trabold, B.M.; Novoa, D.; Abdolvand, A.; Russell, P.S.J. Selective excitation of higher order modes in hollow-core PCF via prism-coupling. *Opt. Lett.* **2014**, *39*, 3736. [[CrossRef](#)] [[PubMed](#)]
84. Long, G.L.; Winefordner, J.D. Limit of Detection A Closer Look at the IUPAC Definition. *Anal. Chem.* **1983**, *55*, 712A–724A. [[CrossRef](#)]
85. McCreery, R.L. *Raman Spectroscopy for Chemical Analysis*; John Wiley & Sons: Hoboken, NJ, USA, 2000; ISBN 0471252875.



© 2017 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>).