

**Erratum: Laser-Frequency Stabilization via a Quasimonolithic Mach-Zehnder Interferometer with Arms of Unequal Length and Balanced dc Readout**  
[Phys. Rev. Applied 7, 024027 (2017)]

Oliver Gerberding, Katharina-Sophie Isleif, Moritz Mehmet, Karsten Danzmann, and Gerhard Heinzel



(Received 23 February 2018; published 6 March 2018)

DOI: [10.1103/PhysRevApplied.9.039902](https://doi.org/10.1103/PhysRevApplied.9.039902)

This paper was published online on 24 February 2017 with an error. The paper states an interferometer arm-length difference of 12.5 cm. This value is incorrect. The actual arm-length difference is  $\Delta l \approx 6.25$  cm. As a result, the stated frequency stability goal is modified to 9.0 kHz/ $\sqrt{\text{Hz}}$  instead of the originally stated 4.5 kHz/ $\sqrt{\text{Hz}}$ . None of the frequency measurement results are affected.

---

*Published by the American Physical Society under the terms of the [Creative Commons Attribution 4.0 International](https://creativecommons.org/licenses/by/4.0/) license. Further distribution of this work must maintain attribution to the author(s) and the published articles title, journal citation, and DOI.*