CrossMark click for updates

## OPEN ACCESS

**Citation:** Murakami S, Verdonschot RG, Kakimoto N, Sumida I, Fujiwara M, Ogawa K, et al. (2016) Preventing Complications from High-Dose Rate Brachytherapy when Treating Mobile Tongue Cancer via the Application of a Modular Lead-Lined Spacer. PLoS ONE 11(4): e0154226. doi:10.1371/journal. pone.0154226

**Editor:** Qinghui Zhang, North Shore Long Island Jewish Health System, UNITED STATES

Received: January 21, 2016

Accepted: April 11, 2016

Published: April 29, 2016

**Copyright:** © 2016 Murakami et al. This is an open access article distributed under the terms of the <u>Creative Commons Attribution License</u>, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Data Availability Statement: All relevant data are within the paper (S1 Appendix).

**Funding:** The authors have no support or funding to report.

**Competing Interests:** The authors have declared that no competing interests exist.

**RESEARCH ARTICLE** 

# Preventing Complications from High-Dose Rate Brachytherapy when Treating Mobile Tongue Cancer via the Application of a Modular Lead-Lined Spacer

Shumei Murakami<sup>1,2</sup>, Rinus G. Verdonschot<sup>1</sup>\*, Naoya Kakimoto<sup>1,2</sup>, Iori Sumida<sup>1,2</sup>, Masateru Fujiwara<sup>2</sup>, Kazuhiko Ogawa<sup>2</sup>, Souhei Furukawa<sup>1</sup>

1 Department of Oral and Maxillofacial Radiology, Osaka University Graduate School of Dentistry, Suita, Osaka, Japan, 2 Department of Radiation Oncology, Osaka University, Graduate School of Medicine, Suita, Osaka, Japan

\* rinusverdonschot@gmail.com

## Abstract

## Purpose

To point out the advantages and drawbacks of high-dose rate brachytherapy in the treatment of mobile tongue cancer and indicate the clinical importance of modular lead-lined spacers when applying this technique to patients.

## Methods

First, all basic steps to construct the modular spacer are shown. Second, we simulate and evaluate the dose rate reduction for a wide range of spacer configurations.

## Results

With increasing distance to the source absorbed doses dropped considerably. Significantly more shielding was obtained when lead was added to the spacer and this effect was most pronounced on shorter (i.e. more clinically relevant) distances to the source.

## Conclusions

The modular spacer represents an important addition to the planning and treatment stages of mobile tongue cancer using HDR-ISBT.

## Introduction

According to a report by the World Health Organization (WHO) there is a marked rise in the incidence rates for oral cancers in various countries and regions [1] and it has been projected that in the US alone between 2010 and 2020 yearly about 3.6 to 4.4 billion dollar will be spent

on the treatment of head and neck cancer [2]. Within the head and neck regions, carcinoma of the tongue is one of the most common types of oral cancer [3]. When treating this particular type of cancer interstitial brachytherapy (ISBT) is an important treatment option as it may greatly preserve tongue function and mobility (compared to other treatment options). Within ISBT a division can be made between Low-Dose Rate (LDR) and High-Dose Rate (HDR) treatments. The main difference between both types is that for LDR-ISBT lower dose rates (60 mGy / minute) are applied for a prolonged period (about 6–7 days) whereas HDR-ISBT applies a higher dose rate (1200 mGy / minute) for a much shorter period (10 times lasting about 5 minutes each). It has been estimated that the global market for brachytherapy has reached a value of about US\$ 680 million in 2013, for which both LDR- and HDR-ISBT segments accounted for about 70% [4].

Although LDR-ISBT is commonly applied in the treatment of mobile tongue cancer, HDR-ISBT does have certain advantages. For instance, the treatment duration for each session is drastically reduced (whereas LDR-ISBT patients have to wear the radioactive implants during the whole treatment period). Another advantage of HDR-ISBT is that due to the fact that it employs fractionated irradiation cancer cells are damaged, however, normal cells are more preserved compared to treatments applying a continuous irradiation regime. Additionally, the amount of radiation exposed to the medical personnel for LDR-ISBT (e.g., due to initial insertion and periodical inspection of the implants) can be avoided. However, one concern with HDR-ISBT is that although it has the advantage of delivering a high dose locally to a specific target for a short time, the to-be-irradiated areas are usually in close proximity to the gingiva and the mandible (as tongue carcinomas typically arise at the lateral border of the tongue). These regions have been shown to be sensitive to the dose rate of radiation even up to the point that severe complications such as osteoradionecrosis or osteomyelitis may occur [5]. For instance, it has been shown that patients receiving a total dose of more than 50-60 Gy have a greater chance of developing mandibular bone complications [6]. Additionally, it has been reported that these complications are even more prevalent when EBRT (external beam radiotherapy) is applied as a combination treatment [7,8]. Another study indicated that the incidence rate for osteoradionecrosis is relatively high. About 8.2% of patients taken from a large sample (N = 830) documenting a 30-year period (i.e. between 1969 and 1999) displayed osteoradionecrosis to a variety of degrees [9]. Additionally, a recent study [10] indicated a mean weighted prevalence of 5.3% for osteoradionecrosis following brachytherapy based on a metaanalysis of several studies published between 1990-2008.

To prevent severe complications such as osteoradionecrosis, which has been reported to be extremely painful for patients [7,11], the dose to normal tissue needs to be attenuated as much as possible. This can be achieved by increasing the distance between the irradiation source and the mandible by means of a "spacer" (a prosthesis typically made of resin or silicon impression material placed between the mandible and the tongue). It has been shown that when the thickness of the spacer is increased to >10 mm the radiation dose can be reduced by about 65% or more for both radium and iridium treatment regimes using LDR-ISBT [7,12]. Additionally, Obinata et al. [13] in a clinical study using LDR-ISBT found that 18.2% of patients having a spacer less than 10 mm thick developed osteoradionecrosis whereas when the spacer was minimally 10 mm thick typically no osteoradionecrosis occurred. In all, it seems reasonable to assume that the thicker the spacer the more radiation attenuation will be obtained.

However, the space between the tongue and the mandible is quite limited (also depending on tumor size) and in several experimental and clinical studies using low dose-rate radioiso-topes [14-16] it has been suggested that spacers using embedded metals known for their radiation shielding properties (i.e. having a high atomic number) should be beneficial to further reduce irradiation exposure without increasing the thickness to impractical proportions.

However, to date the most prevalent spacer type in treating tongue cancer with interstitial brachytherapy remains to be the unshielded acrylic resin spacer (e.g. [17-19]) and to the best of our knowledge, there has been no report assessing the importance of modular spacers to be used in HDR-ISBT.

The aim of this paper is therefore to show how modular spacers with metal shielding can be readily constructed from suitable dental materials and to point out their clinical importance. Furthermore, we demonstrate how modular spacers can be used in a medical setting, starting with the treatment planning (using 3D-CT), how metal shielding can be applied after the planning stage, and how these spacers can be applied during the treatment stage for mobile tongue cancer using HDR-ISBT.

## **Materials and Methods**

The construction of the modular spacer consists of several steps depicted in Fig 1. First, a teeth arch impression is taken using an alginate impression material (COE Alginate; GC America Inc., Chicago, IL, USA) commonly used in dentistry (B-C; 1 minute) and a plaster model is constructed from the arch impression (D-F; 1 hour).

Subsequently, sheet-type light photo-polymerization cured resin (Splint-Resin LC; GC Corporation Ltd, Tokyo, Japan; effective atomic number of 12.5 with a density of 1.2 g/cm3) commonly used in dentistry is added to the plaster model (G; i.e. tooth part; 1 minute) and is hardened with a light cure (H; 5 minutes). The next step is to add resin to the lateral part of the spacer (1 minute) with a subsequent light cure (5 minutes), which will form the barrier between the tongue and the mandible (I-J). Next, a groove will be drilled in the lateral part (K-L; 5–10 minutes) that later will hold the shielding material (e.g. lead with an atomic number of 82 and a density of 9.8 g/cm3). At this point the spacer is ready to be used for 3D CT planning. Once the planning is finished the required amount of shielding material will be inserted and resin will be added to the top with subsequent light curing (M-N; 5 minutes) which will result in the final product (O). If the spacer requires additional adjustment, resin (and/or shielding) can be easily added or removed from the spacer during the actual planning or treatment session. In case the treating physician opts to order the modular spacer from a dental workshop (instead of producing it on-site) it is still possible and easy to perform the last steps (M-O; Fig 2) on-site instead of sending the spacer back to the workshop (which usually takes a long time) to seal the groove holding the shielding material (see Fig 2).

## Dose Rate Reduction Assessment for the Modular Spacer According to TG-43

To evaluate the dose rate reduction for a wide range of potential configurations we performed a number of computer simulations. We first arranged a HDR radiation source (Iridium-192) in either a single or double plane (for large tumors). Following Tsai et al. (6) and Kakimoto et al. [20] who reported that 50–60 Gy would be the maximum dose, we set 60 Gy at 5 mm offset from the lateral source. We simulated the absorbed dose at 10, 15, 20, 25, 30, 35, 40 mm distance from the lateral source combined with a varying lead thickness (0–5 mm).

Absorbed dose calculations are based on the mathematical formulas described in the protocol "Task Group No. 43 for Ir-192 sources" (TG-43) by the American Association of Physicists in Medicine [21,22]. The dwell weights of all sources were augmented by geometric optimization [23] to make the dose distribution homogeneous. The dwell times were calculated to achieve the reference point given a prescription dose of 60 Gy. In terms of the attenuated dose through lead, a half-value-layer of 3 mm was used for the lead material [24]. Other regions except for the lead were assumed to be in water when calculating the absorbed dose.





Fig 1. Construction steps of a modular spacer containing shielding material.

doi:10.1371/journal.pone.0154226.g001

#### Results

As expected we found that when lead was added significantly more shielding was obtained and absorbed doses dropped dramatically (see Fig 3; right pane). This effect was most pronounced on shorter distances to the source. For example, for the shortest distance between the mandible and the reference point (e.g. 10 mm from the source or 5 mm from the 60 Gy reference point) the dose reduction can go from 51.2% (30.7 Gy) of the 60 Gy dosage at the reference point for



Fig 2. The modular spacer: inserting and holding the lead shield.

doi:10.1371/journal.pone.0154226.g002

Pb-0 (indicating a 5 mm modular spacer with no lead shielding) to 40.7% (24.4 Gy) if 1 mm lead shielding would have been applied with 4 mm resin spacer (2 mm resin on both sides).

# Dose Rate Reduction Assessment Based on Monte-Carlo (MC) Simulations

To further validate the TG-43 results we also implemented a Monte-Carlo simulation [25] to estimate the simulated photon energy absorption at each measurement point using the same configurations. In our simulations, data pertaining to the specific photon interaction cross-section were derived from [26,27], the data of the relative probability and the energy of generating gamma-rays from Ir-192 were refereed to a radioisotope database [28] and the number of generated gamma-rays (histories) was ~1.0 x 10^8. The results of the MC simulations can be seen in Fig.3 (left pane) as well as in Table C in the <u>S1 Appendix</u>. A Welch two-sample t-test indicated that they were not statistically different from the earlier TG-43 simulations, t(152.3) = 1.29, p = 0.2.

## Discussion

When treating tongue carcinoma, high dose-rate interstitial brachytherapy (ISBT) is an important treatment option. Typically, ISBT administers a dose of a particular strength locally to a to-be-irradiated target. However, the mandible and gingiva due to their close proximity to the irradiation site may inadvertently also receive some part of this dose. This could lead to severe complications such as osteomyelitis and even osteoradionecrosis. This paper reports how to construct a modular spacer that can be used for the treatment of tongue carcinoma using HDR-ISBT.

We started out by specifying the basic steps necessary to construct the modular spacer and subsequently assessed its effectiveness in attenuating the dosage to the mandible. The results suggest that this type of spacer significantly reduces the dosage received by the mandible and allows for a large variety of configurations to be used in a clinical setting. Constructing a spacer





Fig 3. Percentage doses for various spacer types (Pb in mm) and distances (in mm) split out by plane and simulation type (Monte-Carlo vs. TG-43).

doi:10.1371/journal.pone.0154226.g003

in a modular way is not only quick (total construction time is less than 2 hours) but also allows physicians to rapidly adjust to a wide range of clinical conditions by manipulating the range of the thickness of the spacer itself, the thickness of the resin and the thickness of the lead shield. Take for instance an extreme situation in which there is almost no space between the mandible and the tongue due to excessive tumor size. In such extreme cases it is often opted to not implement a spacer. However, with a modular spacer it is technically possible to create a small 5 mm spacer with 2 mm resin on each side holding a 1 mm lead shield in its groove. Therefore, a small modular spacer with a 1mm lead shield would be strongly recommended, as it will greatly attenuate the absorbed radiation dose by the mandible compared to a spacer-less situation.

The effectiveness of adding metal shielding to clinical spacers was investigated in 1994 by Fujita et al. [14] who by using Lipowitz-metal (2 mm) embedded in a fixed spacer obtained marked dose reduction effects between metal lined- and unlined resin spacers in an experimental and clinical study that used low dose rate radioisotopes. Although for both radium and iridium irradiation sources there were significant dose reductions, the effect was most prevalent for iridium most likely due its weak gamma ray energy emissions [12,14].

One difference between Fujita et al. and the current study is that their spacer is likely intended to be used in LDR-ISBT in which exact 3D CT planning is not as central as in HDR-ISBT (as the radioisotope typically cannot be exactly set in the planned location). Additionally, their spacer involves the use of Lipowitz-metal (a.k.a. Wood's metal) which has the advantage of a relatively low melting point (70C) making it easy to handle as a melting facility is not required. However, Lipowitz-metal also has certain significant disadvantages such as the fact that it contains cadmium which is extremely toxic (even in trace quantities). Another advantage to use lead is its higher atomic number which delivers higher protection against irradiation. Lead is also cheaper and softer than Lipowitz-metal and as most radiation treatment facilities have a lead-melting device we opted to use lead in combination with our modular spacer.

It is also important to point out that in Fujita et al. (14], p.592) it can be seen that relative dose reduction rates (DRR) are not uniform for each patient (some patients' DRR diverge by more than 15%). This indicates that some patients may benefit from additional shielding whereas standard shielding may be sufficient for others. These data were acquired using Thermoluminescent Dosimetry rods (TLD rods) to measure the absorbed dose and the measures reported in the current paper fit well with them. For instance, when comparing an unshielded resin spacer on 15 mm from the source (i.e. 10 mm from lateral reference point) the value dropped from 60 Gy to 20.1 Gy (~66.5% reduction) in our data. This closely matches the ~66% dose reduction rate reported in Fig.3 of Fujita et al. ([14]; p. 591) using actual TLD rods. Similarly, when adding 2 mm of lead shielding at this distance the absorbed dose dropped to 12.7 Gy (~79% dose reduction) in our simulation corresponding to the ~75% dose reduction rate of Fujita et al. [14] obtained (as they used another shielding material, i.e. Lipowitz metal under different conditions, the number is not exactly similar). See Tables A-C in the S1 Appendix (and Fig 3) for all the information containing our absorbed dose calculations, their reductions and the effects of lead and single or double plane sources (note that double-plane sources distribute the dose over a larger surface and have therefore overall lower values than single-plane sources).

Although this conclusion is clear, it is surprising that few papers dealing with ISBT and carcinoma of the tongue reported to have used lead-lined modular spacers. Most reports on LDR-ISBT involve a resin-only [17] or lead-only spacer (e.g. [29,30]) and this pattern is similar for HDR-ISBT (e.g. [31,32]).

We believe that by using a modular spacer there is much to be gained both in the planning and treatment stages. For instance, a 10 mm resin only spacer, although providing satisfactory protection to a certain extent [13], is clearly surpassed by a lead-lined modular spacer in terms of the absorbed dose by the mandible (see our estimations in the results section and Tables A-C in the <u>S1 Appendix</u>). Additionally, mandible lead shields may produce various side effects of having lead (a toxic compound) in the oral cavity even when it is applied for a limited amount of time. Additionally, the specific nature of the interface between tissue and the surface of lead is important due to potential secondary scatter radiation emitted from the surface of a lead-only spacer. For that reason, it is more favorable to create some distance as present in the current modular spacer (e.g. when enveloping the lead into the resin).

#### Using the modular spacer in clinical practice

Nowadays, treatment planning (especially in HDR-ISBT) relies heavily on 3D CT images, which have become indispensible to assess the individual adjustment and optimization of the dose distribution for each patient's irradiation site [23,33,34]. One point of concern is that metals having high atomic numbers (such as lead or Lipowitz-metal) produce large metal artifacts

on 3D CT images (i.e. due to scatter, beam hardening, and photon starvation; e.g. [35,36]). This poses difficulties for treatment planning if shielding devices such as custom designed mandibular lead shields [29] or non-modular (e.g. ready made) shielded spacers are involved in the planning and treatment. Ideally, the spacer should already be present when using 3D-CT planning, as its size and position should be taken into account to determine the exact dose distribution for the to-be irradiated sites (see Fig 4).



Fig 4. Modular spacer during 3D-CT planning without lead-shield (i.e. eliciting no artifacts).

doi:10.1371/journal.pone.0154226.g004

PLOS ONE

Therefore, using the currently described modular spacer one can first determine the appropriate size of the spacer, its resin thickness and shielding factor according to each specific patient's circumstance. Then create the spacer (e.g. following the protocol provided in this paper), perform pre-treatment 3D CT planning (with the unshielded spacer present), and finally add the necessary amount of shielding material before treatment. In this way all the benefits from precise 3D-CT dosage-distributional planning are preserved whilst keeping sufficient protection of the mandible due to the modular shielding.

The modular spacer is well suited for both low-dose (LDR) and high-dose rate (HDR) ISBT. For instance, in LDR-ISBT patients receive a small fraction of the irradiation dose for a long time (e.g. continuously for over a week). In this situation patients have to wear the spacer constantly, which is quite uncomfortable and detrimental to oral hygiene. Additionally, medical personnel also are exposed to radiation and patients are isolated from their surroundings for a significant amount of time. In this situation typically the thinner the spacer the more comfortable the treatment is for the patient. By adding metal shielding to the modular spacer adequate dose attenuation can be obtained whilst keeping the spacer comfortable. Conversely, for HDR-ISBT the patient will receive a much higher dose for a shorter amount of time. In this case comfort is somewhat less important as the irradiation time is much shorter (e.g. 5 minutes per session). Therefore, a thicker spacer with increased shielding can be opted for to obtain more protection for a short period of time.

#### Study limitation

We have started to implement the modular spacer in 14 patients in our hospital starting from two years ago. Although these patients were observed for a relatively short period of time, both late (~2 year) and acute complications (such as redness, erosion, and ulcers in the gingiva) have not been observed to date.

However, it will take time (5–10 years) to see the actual clinical outcomes of this spacer in terms of its shielding effects from osteoradionecrosis (as this condition usually takes years to develop) but the absorbed dose estimations are very encouraging and closely follow the TLD data from Fujita et al. [14].

#### Conclusion

We believe that a modular spacer (allowing for on-site insertion or removal of metal shielding as well as further customization of the resin) represents a significant improvement over leadonly or resin-only non-modular spacers. Although the positive points of metal-lined spacers were already introduced roughly 20 years ago [13] they are still not in standard practice during the treatment of tongue carcinoma. The benefits nevertheless are very clear: First of all, the modular spacer is not overly complicated to construct and the construction process overall takes less than 2 hours. Secondly, it provides the opportunity for on-site modification without the strict necessity to return it to a dental workshop (which would take time). This becomes important for instance when patients' circumstances unexpectedly change during the treatment. Lastly, it provides important options for the planning (3D-CT) and treatment processes (e.g. wide range of sizes), which are not feasible using resin-only or lead-only spacers. Lastly, modular lead-lined spacers provide excellent protection against particular harmful side effects of ISBT and therefore help to prevent serious complications such as osteomyelitis and osteoradionecrosis.

In conclusion, the modular spacer represents an important addition to the tools available for the planning and treatment of mobile tongue cancer using HDR ISBT and is readily applicable for usage on a wide scale in the treatment of tongue cancer.

## **Supporting Information**

**S1 Appendix.** Table A. TG-43 simulations (in Gy) for single/double plane concerning five lead (Pb) thickness variations at various distances (in mm) from the lateral source. Table B. The effect of lead shielding (based on TG-43). Table C. Monte-Carlo simulations (in Gy) for single/double plane concerning five lead (Pb) thickness variations at various distances (in mm) from the lateral source. (PDF)

## **Author Contributions**

Conceived and designed the experiments: SM RV NK IS MF KO SF. Performed the experiments: SM RV IS NK. Analyzed the data: SM RV NK IS MF KO SF. Contributed reagents/ materials/analysis tools: SM RV IS. Wrote the paper: SM RV NK IS MF KO SF.

#### References

- 1. Petersen PE. Oral cancer prevention and control—the approach of the World Health Organization. Oral Oncol 2009; 45(4–5):454–460. doi: 10.1016/j.oraloncology.2008.05.023 PMID: 18804412
- Mariotto AB, Yabroff KR, Shao Y, Feuer EJ, Brown ML. Projections of the cost of cancer care in the United States: 2010–2020. J Natl Cancer Inst 2011; 103:117–28. doi: <u>10.1093/jnci/djq495</u> PMID: <u>21228314</u>
- Siegel R, Ma J, Zhaohui Z, Jemal A. Cancer statistics, 2014. CA Cancer J Clin. 2014; 64:9–29. doi: <u>10.</u> <u>3322/caac.21208</u> PMID: <u>24399786</u>
- Goethals, P-E & Zimmermann, R. (2014). Opportunities in Brachytherapy (Edition 2014). Publisher: MEDraysntell. <u>http://medraysintell.com/Brachytherapy.html</u>
- Umeda M, Komatsubara H, Nishimatsu N, Yokoo S, Shibuya Y, Komori T. High-dose rate interstitial brachytherapy for stage I–II tongue cancer. Oral Surg Oral Med Oral Pathol Oral Radiol Endod 2000; 90:667–70. PMID: <u>11077395</u>
- Tsai CJ, Hofstede TM, Sturgis EM, Garden AS, Lindberg ME, Wei Q, et al. Osteoradionecrosis and radiation dose to the mandible in patients with oropharyngeal cancer. Int J Radiat Oncol Biol Phys 2013, 85:415–420. doi: 10.1016/j.ijrobp.2012.05.032 PMID: 22795804
- Fujita M, Hirokawa Y, Kashiwado K, Akagi Y, Kashimoto K, Kiriu H, et al. An analysis of mandibular bone complications in radiotherapy for T1 and T2 carcinoma of the oral tongue. Int J Radiat Oncol Biol Phys 1996; 34:333–9. PMID: 8567334
- Wendt CD, Peters LJ, Delclos L, Ang KK, Morrison WH, Maor MH, et al. Primary radiotherapy in the treatment of stage I and II oral tongue cancers: Importance of the proportion of therapy delivered with interstitial therapy. Int J Radiat Oncol Biol Phys 1990; 18:1287–1292. PMID: <u>2370178</u>
- Reuther T, Schuster T, Mende U, Kübler A. Osteonecrosis of the jaws as a side effect of radiotherapy of the head and neck tumour patients—a report of a thirty year retrospective review. Int J Oral Maxillofac Surg. 2003; 32:289–295. PMID: 12767877
- Peterson DE, Doerr W, Hovan A, Pinto A, Saunders D, Elting LS, et al. Osteoradionecrosis in cancer patients: the evidence base for treatment-dependent frequency, current management strategies, and future studies. Support Care Cancer. 2010; 18:1089–98. doi: <u>10.1007/s00520-010-0898-6</u> PMID: <u>20526784</u>
- Kudoh T, Ikushima H, Kudoh K, Tokuyama R, Osaki K, Furutani S, et al. High-dose-rate brachytherapy for patients with maxillary gingival carcinoma using a novel customized intraoral mold technique. Oral Surg Oral Med Oral Pathol Oral Radiol Endod 2010; 109:e102–8. doi: <u>10.1016/j.tripleo.2009.10.019</u> PMID: <u>20123391</u>
- Fujita M, Tamamoto M, Hirokawa Y, Kashiwado K, Akagi Y, Kashimoto K, et al. Experimental and clinical studies on dose reduction effects of spacers in interstitial brachytherapy for carcinoma of the mobile tongue. Oral Surg Oral Med Oral Pathol 1993; 76: 797–803. PMID: <u>8284088</u>
- Obinata K, Ohmori K, Tuchiya K, Nishioka T, Shirato H, Nakamura M. Clinical study of a spacer to help prevent osteoradionecrosis resulting from brachytherapy for tongue cancer. Oral Surg Oral Med Oral Pathol Oral Radiol Endod 2003; 95:246–50. PMID: <u>12582367</u>
- Fujita M, Hirokawa Y, Tamamoto M, Kashiwado K, Akagi Y, Kashimoto K, et al. Dose-reducing effect of Lopowitz metal-embedded spacers in interstitial brachytherapy for carcinoma of the mobile tongue. Oral Surg Oral Med Oral Pathol 1994; 77:589–593. PMID: <u>8065721</u>

- Kakimoto N, Inoue T, Inoue T, Murakami S, Furukawa S, Yoshida K, et al. Results of low- and highdose-rate interstitial brachytherapy for T3 mobile tongue cancer. Radiother Oncol 2003; 68:123–128. PMID: <u>12972306</u>
- Mazeron JJ, Crook JM, Benck V, Marinello G, Martin M, Raynal M, et al. Iridium 192 implantation of T1 and T2 carcinomas of the mobile tongue. Int J Radiat Oncol Biol Phys 1990; 19:1369–1376. PMID: 2262360
- Miura M, Takeda M, Sasaki T, Inoue T, Nakayama T, Fukuda H, et al. Factors affecting mandibular complications in low dose brachytherapy for oral tongue carcinoma with special reference to spacer. Int J Radiat Oncol Biol Phys 1998; 41:763–70. PMID: <u>9652836</u>
- Oota S, Shibuya H, Yoshimura RI, Watanabe H, Miura M. Brachytherapy of stage mobile tongue carcinoma. Prediction of local control and QOL. Radiat Oncol 2006; 1:21. PMID: <u>16834784</u>
- Yoshimura RI, Shibuya H, Hayashi K, Nakagawa K, Toda K, Watanabe H, et al. Repeat brachytherapy for patients with residual or recurrent tumors of oral cavity. Int J Radiat Oncol Biol Phys 2012; 83:1198– 204. doi: <u>10.1016/j.ijrobp.2011.09.018</u> PMID: <u>22099049</u>
- Kakimoto N, Inoue T, Inoue T, Murakami S, Furukawa S, Yoshida K, et al. High-dose-rate interstitial brachytherapy for mobile tongue cancer: Influence of the non-irradiated period. Anticancer Res 2006; 26:3933e3937. PMID: <u>17094424</u>
- Nath R, Anderson LL, Luxton G, Weaver KA, Williamson JF, Meigooni AS, Dosimetry of interstitial brachytherapy sources: Recommendations of the AAPM Radiation Therapy Committee Task Group No. 43. Med Phys 1995; 22:209–234. PMID: <u>7565352</u>
- Rivard MJ, Coursey BM, DeWerd LA, Hanson WF, Huq MS, Ibbott GS, et al. Update of AAPM Task Group No. 43 Report: A revised AAPM protocol for brachytherapy dose calculations. Med Phys 2004; 31:633–674. PMID: <u>15070264</u>
- Edmundson GK. Geometry based optimization for stepping source implants. In: Martinez AA, Orton CG, Mould RF, editors. Brachytherapy HDR and LDR. Columbia, MD: Nucletron, 1990. p. 184–192
- Delacroix D, Guerre JP, Leblanc P, Hickman C, Radiation Protection Dosimetry—Radionuclide and Radiation Protection Data Handbook (Kent, England: Nuclear Technology Publishing, 1998), p. 112
- Kato H. Application of the Monte Carlo simulation 1: X-ray image forming process. Japanese Journal of Radiological Technology. 2014 Oct; 70(10):1181–7. PMID: 25327429
- Berger MJ, Hubbell JH, Seltzer SM, Chang J, Coursey JS, Sukumar R, et al. XCOM: Photon cross sections database (NBSIR 87–3597), NIST. 1998.
- Hubbell JH, Veigele WJ, Briggs EA, Brown RT, Cromer DT, Howerton RJ. Atomic form factors, incoherent scattering functions, and photon cross sections. J Phys Chem Ref Data 1975; 4(3): 471–538
- Japan Radioisotope Association. Radioisotope Pocket Data Book 11th Edition. Tokyo, Maruzen, 2011. ISBN 978-4-89073-211-1.
- Goineau A, Piot B, Malard O, Ferron C, Lisbona A, Cassagnau E, et al. Postoperative interstitial brachytherapy for resectable squamous cell carcinoma of the tongue. Brachytherapy 2015; 14:71–76. doi: <u>10.</u> <u>1016/j.brachy.2014.09.011</u> PMID: <u>25454575</u>
- Mazeron JJ, Ardiet JM, Haie-Méder C, Kovács G, Levendag P, Peiffert D, et al. GEC-ESTRO recommendations for brachytherapy for head and neck squamous cell carcinomas. Radiother Oncol. 2009; 91:150–6. doi: 10.1016/j.radonc.2009.01.005 PMID: 19329209
- Lau HY, Hay JH, Flores AD, Threlfall WJ. Seven fractions of twice daily high dose-rate brachytherapy for node-negative carcinoma of the mobile tongue results in loss of therapeutic ratio. Radiother Oncol 1996; 39:15–18. PMID: <u>8735489</u>
- Leung TW, Wong VY, Wong CM, Tung SY, Lui CM, Leung LC, et al. High dose rate brachytherapy for carcinoma of the oral tongue. Int J Radiat Oncol Biol Phys 1997; 39:1113–1120. PMID: <u>9392552</u>
- Lachance B, Béliveau-Nadeau D, Lessard É, Chretien MA, Hsu IC, Pouliot J, et al. Early clinical experience with anatomy-based inverse planning dose optimization for HDR boost of the prostate. Int J Radiat Oncol Biol Phys 2002; 54:86–100. PMID: 12182978
- Sumida I, Shiomi H, Yoshioka Y, Inoue T, Lessard E, Hsu IC, et al. Optimization of dose distribution for HDR brachytherapy of the prostate using attraction-repulsion model. Int J Radiat Oncol Biol Phys 2006; 64:643–649 PMID: 16289907
- Barrett JF, Keat N. Artifacts in CT: Recognition and avoidance. Radiographics 2004; 24:1679–1691. PMID: <u>15537976</u>
- Williamson JF, Whiting BR, Benac J, Murphy RJ, Blaine GJ, O'Sullivan JA, et al. Prospects for Quantitative Computed Tomography Imaging in the Presence of Foreign Metal Bodies Using Statistical Image Reconstruction. Med Phys 2002; 29:2404–2418. PMID: <u>12408315</u>